

LB117895


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35582



A= 0.0000e+000

B= 2.8265e-004

C= 8.3002e-003

Rho= 0.9999150

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.016	0.016	27	0.000	27					1
0.05	0.050	0.053	0.003	158	0.0 %	158	0				6
0.2	0.200	0.197	-0.003	668	0.0 %	668					-2
2.5	2.500	2.500	-0.000	8815	0.0 %	8815					0
5.0	5.000	5.024	0.024	17746	0.0 %	17746					0
7.5	7.500	7.394	-0.106	26130	0.0 %	26130					-1
10.0	10.000	10.066	0.066	35582	0.0 %	35582					1

LB117895 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Dec 2021 12:33:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	245.1	PPB	20 Dec 2021 13:41:06	S	Std
0.05 - 1	50.05	158	-	0.0500	245.1	PPB	20 Dec 2021 13:43:23	S	Std
0.2 - 1	50.2	668	-	0.2000	245.1	PPB	20 Dec 2021 13:45:40	S	Std
2.5 - 1	52.5	8815	-	2.5000	245.1	PPB	20 Dec 2021 13:47:57	S	Std
5.0 - 1	55	17746	-	5.0000	245.1	PPB	20 Dec 2021 13:50:14	S	Std
7.5 - 1	57.5	26130	-	7.5000	245.1	PPB	20 Dec 2021 13:52:29	S	Std
10.0 - 1	510	35582	-	10.0000	245.1	PPB	20 Dec 2021 13:54:45	S	Std
ICV19 - 1	ICV19	14747	4.1766	-245.1		PPB	20 Dec 2021 14:00:04	U	SMPL
ICB19 - 1	ICB19	-63	-0.0095	-245.1		PPB	20 Dec 2021 14:02:18	U	SMPL
CCV72 - 1	CCV72	17343	4.9104	-245.1		PPB	20 Dec 2021 14:04:33	U	SMPL
CCB72 - 1	CCB72	-64	-0.0098	-245.1		PPB	20 Dec 2021 14:06:49	U	SMPL
CRA - 1	CRA	160	0.0535	-245.1		PPB	20 Dec 2021 14:09:03	U	SMPL
HighStd - 1	HighStd	34385	9.7274	-245.1		PPB	20 Dec 2021 14:11:19	U	SMPL
ChkStd - 1	ChkStd	24314	6.8808	-245.1		PPB	20 Dec 2021 14:13:34	U	SMPL
PB141579BL - 1	PBW	-26	0.0010	-245.1		PPB	20 Dec 2021 14:15:50	U	SMPL
PB141579BS - 1	LCSW	13171	3.7311	-245.1		PPB	20 Dec 2021 14:20:55	U	SMPL
M5061-02 - 1	EFF-WW	9	0.0108	-245.1		PPB	20 Dec 2021 14:23:11	U	SMPL
M5151-02 - 1	EFFLUENT-METALS	1426	0.4114	-245.1		PPB	20 Dec 2021 14:25:28	U	SMPL
M5151-02DUP - 1	EFFLUENT-METALSDUP	1335	0.3856	-245.1		PPB	20 Dec 2021 14:27:45	U	SMPL
M5151-02MS - 1	EFFLUENT-METALSMS	13143	3.7232	-245.1		PPB	20 Dec 2021 14:32:36	U	SMPL
M5151-02MSD - 1	EFFLUENT-METALSMSD	14183	4.0172	-245.1		PPB	20 Dec 2021 14:36:23	U	SMPL
CCV73 - 1	CCV73	17040	4.8247	-245.1		PPB	20 Dec 2021 14:38:38	U	SMPL
CCB73 - 1	CCB73	-82	-0.0149	-245.1		PPB	20 Dec 2021 14:40:53	U	SMPL
M5151-02LX5 - 1		374	0.1140	-245.1		PPB	20 Dec 2021 14:43:09	U	SMPL
M5151-02A - 1		17425	4.9336	-245.1		PPB	20 Dec 2021 14:45:24	U	SMPL
CCV74 - 1	CCV74	16964	4.8033	-245.1		PPB	20 Dec 2021 14:47:40	U	SMPL
CCB74 - 1	CCB74	-86	-0.0160	-245.1		PPB	20 Dec 2021 14:49:56	U	SMPL