

LB117909


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37867



A= 0.0000e+000

B= 2.6358e-004

C= 1.3883e-002

Rho= 0.9999909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.016	0.016	9	0.000	9					1
0.2	0.200	0.203	0.003	716	0.0 %	716					2
2.5	2.500	2.470	-0.030	9319	0.0 %	9319					-1
5.0	5.000	4.999	-0.001	18912	0.0 %	18912					0
7.5	7.500	7.517	0.017	28467	0.0 %	28467					0
10.0	10.000	9.995	-0.005	37867	0.0 %	37867					0

LB117909 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Dec 2021 09:23:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000	7470A	PPB	21 Dec 2021 09:33:27	S	Std
0.2 - 1	50.2	716	-	0.2000	7470A	PPB	21 Dec 2021 09:35:44	S	Std
2.5 - 1	52.5	9319	-	2.5000	7470A	PPB	21 Dec 2021 09:38:02	S	Std
5.0 - 1	55	18912	-	5.0000	7470A	PPB	21 Dec 2021 09:42:57	S	Std
7.5 - 1	57.5	28467	-	7.5000	7470A	PPB	21 Dec 2021 09:45:13	S	Std
10.0 - 1	510	37867	-	10.0000	7470A	PPB	21 Dec 2021 09:47:29	S	Std
ICV20 - 1	ICV20	13614	3.6023	-	7470A	PPB	21 Dec 2021 09:50:12	U	SMPL
ICB20 - 1	ICB20	-83	-0.0080	-	7470A	PPB	21 Dec 2021 09:52:27	U	SMPL
CCV75 - 1	CCV75	19040	5.0325	-	7470A	PPB	21 Dec 2021 09:54:42	U	SMPL
CCB75 - 1	CCB75	-91	-0.0101	-	7470A	PPB	21 Dec 2021 09:56:57	U	SMPL
CRA - 1	CRA	713	0.2018	-	7470A	PPB	21 Dec 2021 09:59:12	U	SMPL
HighStd - 1	HighStd	37223	9.8252	-	7470A	PPB	21 Dec 2021 10:01:27	U	SMPL
ChkStd - 1	ChkStd	26134	6.9023	-	7470A	PPB	21 Dec 2021 10:03:42	U	SMPL
PB141594BL - 1	PBW	-95	-0.0112	-	7470A	PPB	21 Dec 2021 10:05:59	U	SMPL
PB141594BS - 1	LCSW	14624	3.8685	-	7470A	PPB	21 Dec 2021 10:08:16	U	SMPL
M5142-02 - 1	A2	148	0.0529	-	7470A	PPB	21 Dec 2021 10:10:32	U	SMPL
M5142-02DUP - 1	A2DUP	221	0.0721	-	7470A	PPB	21 Dec 2021 10:12:48	U	SMPL
M5142-02MS - 1	A2MS	14021	3.7096	-	7470A	PPB	21 Dec 2021 10:15:05	U	SMPL
M5142-02MSD - 1	A2MSD	14476	3.8295	-	7470A	PPB	21 Dec 2021 10:17:22	U	SMPL
PB141595BL - 1	PBW	-59	-0.0017	-	7470A	PPB	21 Dec 2021 10:19:37	U	SMPL
CCV76 - 1	CCV76	19114	5.0520	-	7470A	PPB	21 Dec 2021 10:21:53	U	SMPL
CCB76 - 1	CCB76	-79	-0.0069	-	7470A	PPB	21 Dec 2021 10:24:08	U	SMPL
PB141595BS - 1	LCSW	13796	3.6503	-	7470A	PPB	21 Dec 2021 10:26:23	U	SMPL
M5114-03 - 1	HS-OILY-WATER	232	0.0750	-	7470A	PPB	21 Dec 2021 10:28:38	U	SMPL
M5165-01 - 1	EXCAVATION-INFLUENT	2	0.0144	-	7470A	PPB	21 Dec 2021 10:30:54	U	SMPL
M5165-01DUP - 1	EXCAVATION-INFLUENTDUP	38	0.0239	-	7470A	PPB	21 Dec 2021 10:33:10	U	SMPL
M5165-01MS - 1	EXCAVATION-INFLUENTMS	13634	3.6076	-	7470A	PPB	21 Dec 2021 10:35:26	U	SMPL
M5165-01MSD - 1	EXCAVATION-INFLUENTMSD	12823	3.3938	-	7470A	PPB	21 Dec 2021 10:37:43	U	SMPL
M5165-02 - 1	EXCAVATION-EFFLUENT	-44	0.0023	-	7470A	PPB	21 Dec 2021 10:39:59	U	SMPL
CCV77 - 1	CCV77	19127	5.0554	-	7470A	PPB	21 Dec 2021 10:42:16	U	SMPL
CCB77 - 1	CCB77	-63	-0.0027	-	7470A	PPB	21 Dec 2021 10:44:33	U	SMPL
PB141594TB - 1	PB141539TB	-24	0.0076	-	7470A	PPB	21 Dec 2021 10:46:49	U	SMPL
M5142-02LX5 - 1		43	0.0252	-	7470A	PPB	21 Dec 2021 10:49:04	U	SMPL
M5142-02A - 1		15089	3.9911	-	7470A	PPB	21 Dec 2021 10:51:19	U	SMPL
M5165-01LX5 - 1		-54	-0.0003	-	7470A	PPB	21 Dec 2021 10:53:35	U	SMPL
M5165-01A - 1		16335	4.3195	-	7470A	PPB	21 Dec 2021 10:55:50	U	SMPL
CCV78 - 1	CCV78	18734	4.9518	-	7470A	PPB	21 Dec 2021 10:58:06	U	SMPL
CCB78 - 1	CCB78	-86	-0.0088	-	7470A	PPB	21 Dec 2021 11:00:21	U	SMPL