



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117914

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	12/21/2021	12:15
2	CAL2	1	Water	NA	NA	20.4	7.00	12/21/2021	12:16
3	CAL3	1	Water	NA	NA	20.3	10.03	12/21/2021	12:17
4	ICV	1	Water	NA	NA	20.3	7.02	12/21/2021	12:20
5	CCV1	1	Water	NA	NA	20.3	2.01	12/21/2021	12:22
6	M5192-01	1	Water	NA	NA	20.1	4.97	12/21/2021	12:25
7	M5192-01DUP	1	Water	NA	NA	20.2	4.98	12/21/2021	12:26
8	CCV2	1	Water	NA	NA	20.4	12.03	12/21/2021	12:30

117914

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m5192 WorkList ID : 155416 Department : Wet-Chemistry Date : 12-21-2021 12:01:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/13/2021	Water	M5192-01	pH	Cool 4 deg C	STER15	H11	POOL-WATER	12/13/2021	9040C

Date/Time 12/21/21 12:10
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)

Date/Time 12/21/21 13:00
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)