

LB117919


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

4886E



A= 0.0000e+000

B= 2.0940e-004

C= -1.2110e-001

Rho= 0.9988781

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	-0.103	-0.103	88	0.000	88					-16
0.2	0.200	0.169	-0.031	1384	0.0 %	1384					11
2.5	2.500	2.776	0.276	13834	0.0 %	13834					11
5.0	5.000	5.033	0.033	24614	0.0 %	24614					11
7.5	7.500	7.213	-0.287	35027	0.0 %	35027					11
10.0	10.000	10.112	0.112	48868	0.0 %	48868					11

12/22/15

LB117919 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 21 Dec 2021 15:28:20				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	88	-	0.0000	7471B	PPB	21 Dec 2021 15:28:57	S Std
0.2 - 1	50.2	1384	-	0.2000	7471B	PPB	21 Dec 2021 15:31:14	S Std
2.5 - 1	52.5	13834	-	2.5000	7471B	PPB	21 Dec 2021 15:38:43	S Std
5.0 - 1	55	24614	-	5.0000	7471B	PPB	21 Dec 2021 15:41:00	S Std
7.5 - 1	57.5	35027	-	7.5000	7471B	PPB	21 Dec 2021 15:43:16	S Std
10.0 - 1	58	48868	-	10.0000	7471B	PPB	21 Dec 2021 15:50:49	S Std
ICV21 - 1	ICV21	17836	3.6137	-	7471B	PPB	21 Dec 2021 15:53:54	U SMPL
ICB21 - 1	ICB21	196	-0.0801	-	7471B	PPB	21 Dec 2021 15:56:09	U SMPL
CCV79 - 1	CCV79	24054	4.9158	-	7471B	PPB	21 Dec 2021 15:58:24	U SMPL
CCB79 - 1	CCB79	72	-0.1060	-	7471B	PPB	21 Dec 2021 16:00:39	U SMPL
CRA - 1	CRA	1292	0.1494	-	7471B	PPB	21 Dec 2021 16:09:24	U SMPL
HighStd - 1	HighStd	49468	10.2374	-	7471B	PPB	21 Dec 2021 16:11:39	U SMPL
ChkStd - 1	ChkStd	33193	6.8294	-	7471B	PPB	21 Dec 2021 16:13:55	U SMPL
PB141614BL - 1	PBS	161	-0.0874	-	7471B	PPB	21 Dec 2021 16:16:13	U SMPL
PB141614BS - 1	LCSS	18123	3.6738	-	7471B	PPB	21 Dec 2021 16:21:25	U SMPL
M5178-01 - 1	CL-02-122021	2380	0.3773	-	7471B	PPB	21 Dec 2021 16:23:41	U SMPL
M5179-01 - 1	POST-EXCAVATION-BOTTOM-WEST	506	-0.0151	-	7471B	PPB	21 Dec 2021 16:25:57	U SMPL
M5179-02 - 1	POST-EXCAVATION-BOTTOM-EAST	596	0.0037	-	7471B	PPB	21 Dec 2021 16:28:14	U SMPL
M5179-02DUP - 1	POST-EXCAVATION-BOTTOM-EASTDUP	652	0.0154	-	7471B	PPB	21 Dec 2021 16:30:32	U SMPL
M5179-03 - 1	M5179-02MS	19196	3.8985	-	7471B	PPB	21 Dec 2021 16:35:18	U SMPL
CCV80 - 1	CCV80	24139	4.9336	-	7471B	PPB	21 Dec 2021 16:37:33	U SMPL
CCB80 - 1	CCB80	-34	-0.1282	-	7471B	PPB	21 Dec 2021 16:39:48	U SMPL
M5179-04 - 1	M5179-02MSD	19499	3.9619	-	7471B	PPB	21 Dec 2021 16:42:04	U SMPL
M5179-05 - 1	POST-EXCAVATION-BOTTOM-1	950	0.0778	-	7471B	PPB	21 Dec 2021 16:44:19	U SMPL
M5182-02 - 1	BLIND-DUPLICATE	704	0.0263	-	7471B	PPB	21 Dec 2021 16:46:35	U SMPL
M5182-04 - 1	HB-3-(10-15)-12-17-21	71028	14.7520	-	7471B	PPB	21 Dec 2021 16:48:50	U SMPL
M5182-05 - 1	HB-2-(20-25)-12-17-21	528	-0.0105	-	7471B	PPB	21 Dec 2021 16:51:06	U SMPL
M5182-06 - 1	M5182-05MS	18360	3.7234	-	7471B	PPB	21 Dec 2021 16:55:57	U SMPL
M5182-07 - 1	M5182-05MSD	18458	3.7440	-	7471B	PPB	21 Dec 2021 16:58:13	U SMPL
M5182-08 - 1	HB-1-(18.5-23.5)-12-17-21	722	0.0301	-	7471B	PPB	21 Dec 2021 17:00:30	U SMPL
M5182-10 - 1	HB-3-(20-25)-12-17-21	1436	0.1796	-	7471B	PPB	21 Dec 2021 17:02:47	U SMPL
CCV81 - 1	CCV81	24526	5.0146	-	7471B	PPB	21 Dec 2021 17:05:02	U SMPL
CCB81 - 1	CCB81	-5	-0.1221	-	7471B	PPB	21 Dec 2021 17:07:18	U SMPL
M5179-02LX5 - 1		644	0.0138	-	7471B	PPB	21 Dec 2021 17:09:34	U SMPL
M5179-02A - 1		19355	3.9318	-	7471B	PPB	21 Dec 2021 17:11:49	U SMPL
M5182-04DLX5 - 1		14436	2.9018	-	7471B	PPB	21 Dec 2021 17:14:05	U SMPL
CCV82 - 1	CCV82	24501	5.0094	-	7471B	PPB	21 Dec 2021 17:16:20	U SMPL
CCB82 - 1	CCB82	155	-0.0886	-	7471B	PPB	21 Dec 2021 17:18:36	U SMPL