

LB117923


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51638



A= 0.0000e+000

B= 1.8995e-004

C= -2.4167e-002

Rho= 0.9991369

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.004	-0.004	106	0.000	106					8
0.2	0.200	0.216	0.016	1263	0.0 %	1263					-6
2.5	2.500	2.353	-0.147	12512	0.0 %	12512					2
5.0	5.000	5.102	0.102	26987	0.0 %	26987					3
7.5	7.500	7.749	0.249	40923	0.0 %	40923					-2
10.0	10.000	9.785	-0.215	51638	0.0 %	51638					

LB117923 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 22 Dec 2021 09:55:25					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	106	-	0.0000	7471B	PPB	22 Dec 2021 13:49:11	S	Std
0.2 - 1	50.2	1263	-	0.2000	7471B	PPB	22 Dec 2021 13:51:28	S	Std
2.5 - 1	52.5	12512	-	2.5000	7471B	PPB	22 Dec 2021 13:53:45	S	Std
5.0 - 1	55	26987	-	5.0000	7471B	PPB	22 Dec 2021 13:56:02	S	Std
7.5 - 1	57.5	40923	-	7.5000	7471B	PPB	22 Dec 2021 13:58:19	S	Std
10.0 - 1	510	51638	-	10.0000	7471B	PPB	22 Dec 2021 14:00:35	S	Std
ICV22 - 1	ICV22	21743	4.1059	-	7471B	PPB	22 Dec 2021 14:03:19	U	SMPL
ICB22 - 1	ICB22	155	0.0053	-	7471B	PPB	22 Dec 2021 14:05:33	U	SMPL
CCV83 - 1	CCV83	27349	5.1708	-	7471B	PPB	22 Dec 2021 14:07:48	U	SMPL
CCB83 - 1	CCB83	139	0.0022	-	7471B	PPB	22 Dec 2021 14:10:04	U	SMPL
CRA - 1	CRA	1265	0.2161	-	7471B	PPB	22 Dec 2021 14:12:19	U	SMPL
HighStd - 1	HighStd	52027	9.8584	-	7471B	PPB	22 Dec 2021 14:14:34	U	SMPL
ChkStd - 1	ChkStd	38136	7.2198	-	7471B	PPB	22 Dec 2021 14:16:49	U	SMPL
PB141634BL - 1	PBS	286	0.0302	-	7471B	PPB	22 Dec 2021 14:19:06	U	SMPL
M5187-01 - 1	SASP2-119-1-0-5	5607	1.0409	-	7471B	PPB	22 Dec 2021 14:23:39	U	SMPL
M5187-02 - 1	SASP2-119-1-5-10	1661	0.2913	-	7471B	PPB	22 Dec 2021 14:25:56	U	SMPL
M5187-03 - 1	SASP2-119-1-10-40	582	0.0864	-	7471B	PPB	22 Dec 2021 14:28:13	U	SMPL
M5187-04 - 1	SASP2-119-1-40-57	1850	0.3272	-	7471B	PPB	22 Dec 2021 14:30:29	U	SMPL
M5189-01 - 1	TR-05-122121	2284	0.4097	-	7471B	PPB	22 Dec 2021 14:32:45	U	SMPL
CCV84 - 1	CCV84	27667	5.2312	-	7471B	PPB	22 Dec 2021 14:35:00	U	SMPL
CCB84 - 1	CCB84	122	-0.0010	-	7471B	PPB	22 Dec 2021 14:37:16	U	SMPL
M5190-01 - 1	CHRT-26598	2938	0.5339	-	7471B	PPB	22 Dec 2021 14:39:32	U	SMPL
M5190-03 - 1	RT-880	9664	1.8115	-	7471B	PPB	22 Dec 2021 14:41:47	U	SMPL
M5191-01 - 1	TP-(0-3.5)-12-21-21	6908	1.2880	-	7471B	PPB	22 Dec 2021 14:44:02	U	SMPL
M5191-03 - 1	TP-(3.5-7.0)-12-21-21	4367	0.8053	-	7471B	PPB	22 Dec 2021 14:46:18	U	SMPL
M5198-01 - 1	SB-91SB-22-C2	2469	0.4448	-	7471B	PPB	22 Dec 2021 14:55:38	U	SMPL
M5198-01DUP - 1	SB-91SB-22-C2DUP	2461	0.4433	-	7471B	PPB	22 Dec 2021 14:57:54	U	SMPL
M5198-01MS - 1	SB-91SB-22-C2MS	23603	4.4593	-	7471B	PPB	22 Dec 2021 15:02:42	U	SMPL
M5198-01MSD - 1	SB-91SB-22-C2MSD	23156	4.3743	-	7471B	PPB	22 Dec 2021 15:04:59	U	SMPL
M5198-02 - 1	SB-91SB-22-G2	7180	1.3397	-	7471B	PPB	22 Dec 2021 15:07:16	U	SMPL
M5198-03 - 1	SB-20-C2	666	0.1023	-	7471B	PPB	22 Dec 2021 15:09:32	U	SMPL
CCV85 - 1	CCV85	27800	5.2565	-	7471B	PPB	22 Dec 2021 15:11:47	U	SMPL
CCB85 - 1	CCB85	207	0.0152	-	7471B	PPB	22 Dec 2021 15:14:03	U	SMPL
M5198-04 - 1	SB-20-G2	706	0.1099	-	7471B	PPB	22 Dec 2021 15:16:18	U	SMPL
PB141634BS - 1	LCSS	21927	4.1409	-	7471B	PPB	22 Dec 2021 15:18:34	U	SMPL
M5198-01LX5 - 1		553	0.0809	-	7471B	PPB	22 Dec 2021 15:20:49	U	SMPL
M5198-01A - 1		23763	4.4896	-	7471B	PPB	22 Dec 2021 15:23:05	U	SMPL
CCV86 - 1	CCV86	27892	5.2740	-	7471B	PPB	22 Dec 2021 15:25:21	U	SMPL
CCB86 - 1	CCB86	158	0.0058	-	7471B	PPB	22 Dec 2021 15:27:37	U	SMPL