



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB117925

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP94489
calibration std. phosphate 0.5 ppm	WP94488
calibration std. phosphate 0.3 ppm	WP94487
calibration std. phosphate 0.1 ppm	WP94486
calibration std. phosphate 0.05 ppm	WP94485
calibration std. 0 ppm	WP94484
phosphate CCV std.	WP94491
Combined reagent	WP94493
Phenolphthalein indicator	WP93767
Sodium hydroxide, 1N	WP92662
Phosphate ICV-LCS Std	WP94490

Intercept: -0.0022

Slope: 0.6518

Regression: 0.999938

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		12/22/2021	13:18
2	CAL2	0.05	1	50	50	0.032	0.052	-4	12/22/2021	13:18
3	CAL3	0.10	1	50	50	0.063	0.1	0	12/22/2021	13:19
4	CAL4	0.30	1	50	50	0.188	0.292	2.7	12/22/2021	13:19
5	CAL5	0.50	1	50	50	0.324	0.5	0	12/22/2021	13:20
6	CAL6	1.00	1	50	50	0.651	1.002	-0.2	12/22/2021	13:20



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB117925

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.500	12/22/2021	13:21
2	ICB		1	50	50	0.000	0.003	12/22/2021	13:21
3	CCV1	0.50	1	50	50	0.325	0.502	12/22/2021	13:22
4	CCB1		1	50	50	0.000	0.003	12/22/2021	13:22
5	PB141641BL		1	50	50	0.000	0.003	12/22/2021	13:23
6	PB141641BS	0.50	1	50	50	0.326	0.504	12/22/2021	13:23
7	M5176-03		1	50	50	0.409	0.631	12/22/2021	13:24
8	M5176-03DUP		1	50	50	0.406	0.626	12/22/2021	13:24
9	M5176-03MS	0.50	1	50	50	0.712	1.096	12/22/2021	13:25
10	M5176-03MSD	0.50	1	50	50	0.701	1.079	12/22/2021	13:26
11	M5176-07		1	50	50	0.391	0.603	12/22/2021	13:26
12	M5176-03MS	0.50	2	50	50	0.362	0.559	12/22/2021	13:27
13	M5176-03MSD	0.50	2	50	50	0.351	0.542	12/22/2021	13:27
14	CCV2	0.50	1	50	50	0.326	0.504	12/22/2021	13:28
15	CCB2		1	50	50	0.000	0.003	12/22/2021	13:28