



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117932

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/22/2021	12:30
2	CAL2	1	Water	NA	NA	20.8	7.00	12/22/2021	12:31
3	CAL3	1	Water	NA	NA	20.8	10.03	12/22/2021	12:32
4	ICV	1	Water	NA	NA	20.7	6.99	12/22/2021	12:35
5	CCV1	1	Water	NA	NA	20.6	2.01	12/22/2021	12:36
6	M5208-01	1	Solid	20.04	20	21.3	7.93	12/22/2021	12:48
7	M5208-01DUP	1	Solid	20.03	20	21.4	7.94	12/22/2021	12:49
8	CCV2	1	Water	NA	NA	20.6	12.03	12/22/2021	12:50

YB 117932

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5198 W WorkList ID : 155461 Department : Wet-Chemistry Date : 12-22-2021 14:40:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2021	Solid	M5208-01	pH	Cool 4 deg C	NOBI03	WP-SOIL-1		12/21/2021	9045D

Date/Time 12/22/21 11:20
Raw Sample Received by: JD [signature]
Raw Sample Relinquished by: CF [signature]

Date/Time 12/22/21 6:20
Raw Sample Received by: CP [signature]
Raw Sample Relinquished by: JD [signature]