



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117935

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/22/2021	17:10
2	CAL2	1	Water	NA	NA	20.8	7.00	12/22/2021	17:11
3	CAL3	1	Water	NA	NA	20.7	10.03	12/22/2021	17:12
4	ICV	1	Water	NA	NA	20.6	7.02	12/22/2021	17:15
5	CCV1	1	Water	NA	NA	20.6	2.01	12/22/2021	17:16
6	M5219-02	1	Water	NA	NA	20.6	7.50	12/22/2021	17:20
7	M5219-02DUP	1	Water	NA	NA	20.7	7.51	12/22/2021	17:21
8	CCV2	1	Water	NA	NA	20.6	12.03	12/22/2021	17:22

VB 117935

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m5219 W WorkList ID : 155469 Department : Wet-Chemistry Date : 12-22-2021 16:53:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/22/2021	Water	M5219-02	Corrosivity	Cool 4 deg C	PSEG03	H11	M21013-N48854-1040	12/22/2021	9040C

Date/Time 12/22/21 17:00
Raw Sample Received by: TD (GOC)
Raw Sample Relinquished by: RJ CETA (CUB)

Date/Time 12/22/21 18:00
Raw Sample Received by: RJ CETA (CUB)
Raw Sample Relinquished by: TD (GOC)