

LIB117943

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

12/23/2021 10:53

Test: Total CN

Sample Id	<i>mp/L</i> Result	Dil. 1 +	Response	Errors
ICV1	98.121	0.0	0.083	
ICB1	-0.459	0.0	0.001	
CCV1	241.508	0.0	0.202	
CCB1	-0.983	0.0	0.001	
PB141613BL	-0.400	0.0	0.001	
PB141613BS	98.278	0.0	0.083	
MIDPB141613	244.482	0.0	0.204	
M5182-03	-0.790	0.0	0.001	
CCV2	253.206	0.0	0.212	
CCB2	-0.665	0.0	0.001	

N 10
Mean 93.230
SD 112.8330
CV% 121.03

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 23 09:41:03 2021

Thu Dec 23 10:37:56 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.5743	µg/l	12/23/2021 10:01:28	
5.OPPBCN	A	Total CN	P	4.5936	µg/l	12/23/2021 10:01:29	
10PPBCN	A	Total CN	P	9.5135	µg/l	12/23/2021 10:01:30	
50PPBCN	A	Total CN	P	49.7233	µg/l	12/23/2021 10:01:31	
100PPBCN	A	Total CN	P	102.3433	µg/l	12/23/2021 10:01:32	
250PPBCN	A	Total CN	P	249.6554	µg/l	12/23/2021 10:01:33	
500PPBCN	A	Total CN	P	499.7451	µg/l	12/23/2021 10:01:34	
ICV1	S	Total CN	P	98.1214	µg/l	12/23/2021 10:33:35	
ICB1	S	Total CN	P	-0.4593	µg/l	12/23/2021 10:33:36	
CCV1	S	Total CN	P	241.5079	µg/l	12/23/2021 10:33:39	
CCB1	S	Total CN	P	-0.9829	µg/l	12/23/2021 10:33:41	
PB141613BL	S	Total CN	P	-0.3998	µg/l	12/23/2021 10:33:42	
PB141613BS	S	Total CN	P	98.278	µg/l	12/23/2021 10:33:43	
MIDPB141613	S	Total CN	P	244.4817	µg/l	12/23/2021 10:33:44	
M5182-03	S	Total CN	P	-0.7901	µg/l	12/23/2021 10:37:53	
CCV2	S	Total CN	P	253.2065	µg/l	12/23/2021 10:37:54	
CCB2	S	Total CN	P	-0.6652	µg/l	12/23/2021 10:37:55	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

12/23/2021 10:02


Test Total CN

Accepted 12/23/2021 10:02

Factor 1205
Bias 0.002

Coeff. of det. 0.999968

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.5743	0.0000	-
2	5.0PPBCN	0.005	4.5936	5.0000	-8.1
3	10PPBCN	0.009	9.5135	10.0000	-4.9
4	50PPBCN	0.043	49.7233	50.0000	-0.6
5	100PPBCN	0.087	102.3433	100.0000	2.3
6	250PPBCN	0.209	249.6554	250.0000	-0.1
7	500PPBCN	0.416	499.7451	500.0000	-0.1

12 12/23/21