



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 12/25/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 12/23/2021 16:25 DATE & TIME OUT1: 12/25/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 12/25/2021 08:30 DATE & TIME OUT2: 12/25/2021 10:00Run Number: LB117944BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117944BL	LB117944BL	96.3174	96.3174	100	96.3174	96.3174	96.3174	0.0000	0
2	M5208-03	WP-SOIL-1	89.1521	89.1521	100	89.1613	89.1613	89.1613	0.0092	92
3	M5208-03DUP	WP-SOIL-1DUP	85.0971	85.0971	100	85.1061	85.1061	85.1061	0.0090	90

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WP 117944

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS M5208 WorkList ID : 155517 Department : Wet-Chemistry Date : 12-23-2021 12:10:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/29/2021	Solid	M5208-03	ASTM TS	Cool 4 deg C	NOBI03	H61	WP-SOIL-1	12/21/2021	SM2540 B

Date/Time 12/23/21 15:00
Raw Sample Received by: JD LWC
Raw Sample Relinquished by: JD LWC

Date/Time 12/23/21 16:30
Raw Sample Received by: JD LWC
Raw Sample Relinquished by: JD LWC