

LB 117947


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40980



A= 0.0000e+000

B= 2.4194e-004 slope

C= -4.9251e-002 y-intercept

Rho= 0.9996319

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.012	-0.012	155	0.000	155					1
0.05	0.050					205	0				-6
0.20	0.200	0.188	-0.012	981	0.0 %	981					-4
2.50	2.500	2.406	-0.094	10147	0.0 %	10147					3
5.00	5.000	5.148	0.148	21482	0.0 %	21482					1
7.50	7.500	7.604	0.104	31635	0.0 %	31635					1
10.0	10.000	9.865	-0.135	40980	0.0 %	40980					

LB117947
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	155	-		0 SFAM01.1	PPB	12/23/2021 12:44	Std
	0.2 S01	981	-		0.2 SFAM01.1	PPB	12/23/2021 12:46	Std
	2.5 S02	10147	-		2.5 SFAM01.1	PPB	12/23/2021 12:48	Std
	5 S03	21482	-		5 SFAM01.1	PPB	12/23/2021 12:51	Std
	7.5 S04	31635	-		7.5 SFAM01.1	PPB	12/23/2021 12:55	Std
	10 S05	40980	-		10 SFAM01.1	PPB	12/23/2021 13:00	Std
ICV003	ICV003	16871	4.0325 -		SFAM01.1	PPB	12/23/2021 13:03	SMPL
ICB003	ICB003	91	-0.0272 -		SFAM01.1	PPB	12/23/2021 13:05	SMPL
CCV091	CCV091	21425	5.1343 -		SFAM01.1	PPB	12/23/2021 13:08	SMPL
CCB091	CCB091	85	-0.0287 -		SFAM01.1	PPB	12/23/2021 13:10	SMPL
PB141656BL	PBS656	56	-0.0357 -		SFAM01.1	PPB	12/23/2021 13:12	SMPL
M5035-01	C00U9	1286	0.2619 -		SFAM01.1	PPB	12/23/2021 13:14	SMPL
M5035-02	C00V0	900	0.1685 -		SFAM01.1	PPB	12/23/2021 13:17	SMPL
M5035-03	C00V1	1127	0.2234 -		SFAM01.1	PPB	12/23/2021 13:19	SMPL
M5035-04	C00V2	589	0.0933 -		SFAM01.1	PPB	12/23/2021 13:21	SMPL
M5035-05	C00V2D	670	0.1128 -		SFAM01.1	PPB	12/23/2021 13:23	SMPL
M5035-06	C00V2S	12177	2.8968 -		SFAM01.1	PPB	12/23/2021 13:26	SMPL
PB141657BL	PBS657	145	-0.0142 -		SFAM01.1	PPB	12/23/2021 13:28	SMPL
M5043-01	MH1G62	4469	1.032 -		SFAM01.1	PPB	12/23/2021 13:30	SMPL
M5043-02	MH1G63	38562	9.2804 -		SFAM01.1	PPB	12/23/2021 13:33	SMPL
M5043-03	MH1G64	37421	9.0043 -		SFAM01.1	PPB	12/23/2021 13:35	SMPL
M5043-04	MH1G65	46237	11.1373 -		SFAM01.1	PPB	12/23/2021 13:37	SMPL
M5043-05	MH1G66	132118	31.9152 -		SFAM01.1	PPB	12/23/2021 13:39	SMPL
M5043-06	MH1G67	37754	9.0849 -		SFAM01.1	PPB	12/23/2021 13:42	SMPL
M5043-07	MH1G68	435807	HIGH -		SFAM01.1	PPB	12/23/2021 13:44	
M5043-08	MH1G69	31536	7.5805 -		SFAM01.1	PPB	12/23/2021 13:46	SMPL
M5043-09	MH1G70	124562	30.0871 -		SFAM01.1	PPB	12/23/2021 13:48	SMPL
M5043-10	MH1G71	50338	12.1295 -		SFAM01.1	PPB	12/23/2021 13:50	SMPL
M5043-11	MH1G72	31133	7.483 -		SFAM01.1	PPB	12/23/2021 13:53	SMPL
M5043-12	MH1G73	67224	16.2148 -		SFAM01.1	PPB	12/23/2021 13:55	SMPL
CCV092	CCV092	21930	5.2565 -		SFAM01.1	PPB	12/23/2021 13:57	SMPL
CCB092	CCB092	79	-0.0301 -		SFAM01.1	PPB	12/23/2021 13:59	SMPL
M5043-13	MH1G74	214006	51.7271 -		SFAM01.1	PPB	12/23/2021 14:02	SMPL
M5043-14	MH1G75	106865	25.8055 -		SFAM01.1	PPB	12/23/2021 14:04	SMPL
M5043-15	MH1G76	123456	29.8195 -		SFAM01.1	PPB	12/23/2021 14:06	SMPL
M5043-16	MH1G77	344678	HIGH -		SFAM01.1	PPB	12/23/2021 14:09	
M5043-18	MH1G77D	379701	HIGH -		SFAM01.1	PPB	12/23/2021 14:10	
M5043-19	MH1G77S	408093	HIGH -		SFAM01.1	PPB	12/23/2021 14:12	
M5043-17	MH1G78	257283	62.1975 -		SFAM01.1	PPB	12/23/2021 14:14	SMPL
CCV093	CCV093	21411	5.1309 -		SFAM01.1	PPB	12/23/2021 14:16	SMPL
CCB093	CCB093	38	-0.0401 -		SFAM01.1	PPB	12/23/2021 14:19	SMPL
M5043-04DLX2		23256	5.5773 -		SFAM01.1	PPB	12/23/2021 14:21	SMPL
M5043-05DLX5		26572	6.3795 -		SFAM01.1	PPB	12/23/2021 14:23	SMPL
M5043-07DLX10		68717	16.5761 -		SFAM01.1	PPB	12/23/2021 14:26	SMPL
M5043-09DLX5		24849	5.9627 -		SFAM01.1	PPB	12/23/2021 14:28	SMPL
M5043-10DLX2		24927	5.9816 -		SFAM01.1	PPB	12/23/2021 14:30	SMPL

LB117947

INSTRUMENT ID : CV1

M5043-12DLX5		13391	3.1906 -	SFAM01.1	PPB	12/23/2021 14:32	SMPL
M5043-13DLX10		22766	5.4587 -	SFAM01.1	PPB	12/23/2021 14:35	SMPL
M5043-14DLX15		21694	5.1994 -	SFAM01.1	PPB	12/23/2021 14:37	SMPL
M5043-15DLX5		24819	5.9554 -	SFAM01.1	PPB	12/23/2021 14:39	SMPL
M5043-16DLX10		49265	11.8699 -	SFAM01.1	PPB	12/23/2021 14:41	SMPL
M5043-18DLX10		56496	13.6193 -	SFAM01.1	PPB	12/23/2021 14:44	SMPL
M5043-19DLX10		59803	14.4194 -	SFAM01.1	PPB	12/23/2021 14:46	SMPL
M5043-17DLX10		27670	6.6452 -	SFAM01.1	PPB	12/23/2021 14:48	SMPL
M5043-07DL2X50		13233	3.1523 -	SFAM01.1	PPB	12/23/2021 14:51	SMPL
M5043-16DL2X25		19477	4.663 -	SFAM01.1	PPB	12/23/2021 14:53	SMPL
M5043-18DL2X25		22710	5.4452 -	SFAM01.1	PPB	12/23/2021 14:55	SMPL
M5043-169DL2X25		24671	5.9196 -	SFAM01.1	PPB	12/23/2021 14:57	SMPL
CCV094	CCV094	21768	5.2173 -	SFAM01.1	PPB	12/23/2021 15:00	SMPL
CCB094	CCB094	80	-0.0299 -	SFAM01.1	PPB	12/23/2021 15:02	SMPL