



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117957

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	12/23/2021	13:15
2	M5197-01		1	10	100.00	>212.0	>212.0	12/23/2021	13:40
3	M5197-01DUP		1	10	100.00	>212.0	>212.0	12/23/2021	14:05
4	M5197-03		1	15	100.00	>212.0	>212.0	12/23/2021	14:35
5	M5197-05		1	12	100.00	>212.0	>212.0	12/23/2021	14:58
6	M5219-02		1	14	100.00	>212.0	>212.0	12/23/2021	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 11795

WorkList Name : FP122321 WorkList ID : 155516 Department : Wet-Chemistry Date : 12-23-2021 11:45:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/28/2021	Water	M5197-01	Flash Point	Cool 4 deg C	PSEG03	H61	AUD-1230	12/21/2021	1010B
12/28/2021	Water	M5197-03	Flash Point	Cool 4 deg C	PSEG03	H61	AUD-1231	12/21/2021	1010B
12/28/2021	Water	M5197-05	Flash Point	Cool 4 deg C	PSEG03	H61	AUD-1233-1234	12/21/2021	1010B
12/27/2021	Water	M5219-02	Flash Point	Cool 4 deg C	PSEG03	H11	M21013-N48854-1040	12/22/2021	1010B

Date/Time 12/23/21 12:50
 Raw Sample Received by: KS(wc)
 Raw Sample Relinquished by: RJ(CFA-1040)

Date/Time 12/23/21 15:25
 Raw Sample Received by: RJ(CFA-1040)
 Raw Sample Relinquished by: KS(wc)