



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117960

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/23/2021	15:30
2	CAL2	1	Water	NA	NA	20.6	7.01	12/23/2021	15:31
3	CAL3	1	Water	NA	NA	20.6	10.03	12/23/2021	15:32
4	ICV	1	Water	NA	NA	20.6	7.00	12/23/2021	15:35
5	CCV1	1	Water	NA	NA	20.7	2.01	12/23/2021	15:40
6	M5198-16	1	Solid	20.03	20	20.4	10.88	12/23/2021	15:42
7	M5237-02	1	Solid	20.04	20	21.3	8.76	12/23/2021	15:45
8	M5237-05	1	Solid	20.04	20	21.4	8.34	12/23/2021	15:46
9	M5237-05DUP	1	Solid	20.03	20	21.5	8.35	12/23/2021	15:47
10	CCV2	1	Water	NA	NA	20.7	12.03	12/23/2021	15:50

VB 117960

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m5198 WorkList ID : 15522 Department : Wet-Chemistry Date : 12-23-2021 12:57:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2021	Solid	M5198-16	Corrosivity	Cool 4 deg C	DVIR01		TP-1-5	12/21/2021	9045D
12/23/2021	Solid	M5237-02	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-1S	12/23/2021	9045D
12/23/2021	Solid	M5237-05	Corrosivity	Cool 4 deg C	PSEG03	H11	TP-1D	12/23/2021	9045D

Date/Time 12/23/21 5:00
Raw Sample Received by: 701001
Raw Sample Relinquished by: CP SM

Date/Time 12/23/21 17:30
Raw Sample Received by: CP 500
Raw Sample Relinquished by: 701001