



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/27/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 12/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 12/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117964

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M5225-01	GBC27	1	1.14	8.70	9.84	8.37	83.1	
M5225-02	GBC27D	2	1.14	8.70	9.84	8.37	83.1	
M5225-03	GBC27S	3	1.14	8.70	9.84	8.37	83.1	
M5225-04	GBC28	4	1.13	8.54	9.67	6.59	63.9	
M5225-05	GBC29	5	1.18	8.48	9.66	7.13	70.2	
M5225-06	GBC30	6	1.13	8.79	9.92	7.66	74.3	
M5225-07	GBC31	7	1.15	8.81	9.96	6.88	65.0	
M5225-08	GBC32	8	1.15	8.83	9.98	7.4	70.8	
M5225-09	GBC33	9	1.16	8.72	9.88	7.38	71.3	
M5225-10	GBC34	10	1.19	8.57	9.76	7.41	72.6	
M5225-11	GBC35	11	1.16	8.63	9.79	7.3	71.1	
M5225-12	GBC36	12	1.15	8.79	9.94	8.33	81.7	
M5225-13	GBC37	13	1.11	8.75	9.86	8.1	79.9	
M5225-14	GBC38	14	1.16	8.58	9.74	7.98	79.5	
M5225-15	GBC39	15	1.15	8.82	9.97	8.41	82.3	
M5225-16	GBC40	16	1.17	8.71	9.88	7.89	77.2	
M5225-17	GBC41	17	1.14	8.83	9.97	8.03	78.0	
M5225-18	GBC42	18	1.17	8.73	9.9	7.36	70.9	
M5225-19	GBC43	19	1.14	8.79	9.93	4.31	36.1	
M5225-20	GBC44	20	1.18	8.61	9.79	6.54	62.3	
M5225-21	GBC45	21	1.18	8.78	9.96	6.96	65.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$