


LB117970

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9549e-004

C= 2.7094e-002

Rho= 0.9997373

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	0.037	0.037	52	0.000	52					-
0.2	0.200	0.229	0.029	1031	0.0 %	1031					15
2.5	2.500	2.354	-0.146	11905	0.0 %	11905					-6
5.0	5.000	5.040	0.040	25644	0.0 %	25644					1
7.5	7.500	7.612	0.112	38801	0.0 %	38801					1
10.0	10.000	9.927	-0.073	50643	0.0 %	50643					-)

LB117970 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Dec 2021 09:34:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	52	-	0.0000	7471B	PPB	27 Dec 2021 09:34:44	S	Std
0.2 - 1	50.2	1031	-	0.2000	7471B	PPB	27 Dec 2021 09:37:02	S	Std
2.5 - 1	52.5	11905	-	2.5000	7471B	PPB	27 Dec 2021 09:39:19	S	Std
5.0 - 1	55	25644	-	5.0000	7471B	PPB	27 Dec 2021 09:41:36	S	Std
7.5 - 1	57.5	38801	-	7.5000	7471B	PPB	27 Dec 2021 09:43:52	S	Std
10.0 - 1	511	50643	-	10.0000	7471B	PPB	27 Dec 2021 09:46:09	S	Std
ICV24 - 1	ICV24	19791	3.8960	-	7471B	PPB	27 Dec 2021 09:48:48	U	SMPL
ICB24 - 1	ICB24	-35	0.0203	-	7471B	PPB	27 Dec 2021 09:51:03	U	SMPL
CCV95 - 1	CCV95	25406	4.9937	-	7471B	PPB	27 Dec 2021 09:53:19	U	SMPL
CCB95 - 1	CCB95	-53	0.0167	-	7471B	PPB	27 Dec 2021 09:55:34	U	SMPL
CRA - 1	CRA	1010	0.2245	-	7471B	PPB	27 Dec 2021 09:57:50	U	SMPL
HighStd - 1	HighStd	50501	9.8995	-	7471B	PPB	27 Dec 2021 10:00:05	U	SMPL
ChkStd - 1	ChkStd	36367	7.1364	-	7471B	PPB	27 Dec 2021 10:02:20	U	SMPL
PB141701BL - 1	PBS	-48	0.0177	-	7471B	PPB	27 Dec 2021 10:04:38	U	SMPL
PB141701BS - 1	LCSS	21709	4.2710	-	7471B	PPB	27 Dec 2021 10:06:56	U	SMPL
M5198-10 - 1	TP-1-5	5906	1.1817	-	7471B	PPB	27 Dec 2021 10:09:12	U	SMPL
M5210-01 - 1	124027	512	0.1272	-	7471B	PPB	27 Dec 2021 10:11:29	U	SMPL
M5210-03 - 1	R206	828	0.1890	-	7471B	PPB	27 Dec 2021 10:13:47	U	SMPL
M5210-05 - 1	RT3990	3133	0.6396	-	7471B	PPB	27 Dec 2021 10:16:04	U	SMPL
M5211-01 - 1	ETGI-286	6061	1.2120	-	7471B	PPB	27 Dec 2021 10:18:19	U	SMPL
CCV96 - 1	CCV96	25675	5.0463	-	7471B	PPB	27 Dec 2021 10:20:34	U	SMPL
CCB96 - 1	CCB96	-68	0.0138	-	7471B	PPB	27 Dec 2021 10:22:49	U	SMPL
M5218-01 - 1	SB-01-C2	386	0.1026	-	7471B	PPB	27 Dec 2021 10:25:06	U	SMPL
M5218-01DUP - 1	SB-01-C2DUP	323	0.0902	-	7471B	PPB	27 Dec 2021 10:31:30	U	SMPL
M5218-01MS - 1	SB-01-C2MS	17586	3.4650	-	7471B	PPB	27 Dec 2021 10:36:25	U	SMPL
M5218-01MSD - 1	SB-01-C2MSD	18343	3.6129	-	7471B	PPB	27 Dec 2021 10:38:40	U	SMPL
M5218-02 - 1	SB-01-G2	37	0.0343	-	7471B	PPB	27 Dec 2021 10:40:57	U	SMPL
M5218-03 - 1	SB-21-C2	1177	0.2572	-	7471B	PPB	27 Dec 2021 10:43:13	U	SMPL
M5218-04 - 1	SB-21-G2	2572	0.5299	-	7471B	PPB	27 Dec 2021 10:45:29	U	SMPL
M5224-06 - 1	SB-COMP	16169	3.1880	-	7471B	PPB	27 Dec 2021 10:47:46	U	SMPL
PB141702BL - 1	PBS	-2	0.0267	-	7471B	PPB	27 Dec 2021 10:50:03	U	SMPL
PB141702BS - 1	LCSS	20412	4.0174	-	7471B	PPB	27 Dec 2021 10:52:18	U	SMPL
CCV97 - 1	CCV97	25876	5.0856	-	7471B	PPB	27 Dec 2021 10:54:34	U	SMPL
CCB97 - 1	CCB97	-43	0.0187	-	7471B	PPB	27 Dec 2021 10:56:50	U	SMPL
M5230-01 - 1	PSC124042	3543	0.7197	-	7471B	PPB	27 Dec 2021 10:59:07	U	SMPL
M5230-03 - 1	PSC124055	5429	1.0884	-	7471B	PPB	27 Dec 2021 11:01:23	U	SMPL
M5230-05 - 1	PSC124030	1389	0.2986	-	7471B	PPB	27 Dec 2021 11:03:40	U	SMPL
M5230-07 - 1	PSC164718	788	0.1811	-	7471B	PPB	27 Dec 2021 11:05:55	U	SMPL
M5230-09 - 1	PSC124019	3251	0.6626	-	7471B	PPB	27 Dec 2021 11:08:11	U	SMPL
M5235-01 - 1	HD-02-122321	4611	0.9285	-	7471B	PPB	27 Dec 2021 11:10:28	U	SMPL
M5237-01 - 1	TP-1S	2590	0.5334	-	7471B	PPB	27 Dec 2021 11:12:44	U	SMPL
M5237-04 - 1	TP-1D	760	0.1757	-	7471B	PPB	27 Dec 2021 11:15:01	U	SMPL
M5238-01 - 1	OK-02-122321	1022	0.2269	-	7471B	PPB	27 Dec 2021 11:17:18	U	SMPL
M5239-01 - 1	SASP2-123-1-0-5	2707	0.5563	-	7471B	PPB	27 Dec 2021 11:19:34	U	SMPL
CCV98 - 1	CCV98	26244	5.1575	-	7471B	PPB	27 Dec 2021 11:21:49	U	SMPL
CCB98 - 1	CCB98	-55	0.0163	-	7471B	PPB	27 Dec 2021 11:24:05	U	SMPL
M5239-02 - 1	SASP2-123-1-5-10	11786	2.3311	-	7471B	PPB	27 Dec 2021 11:26:21	U	SMPL
M5239-03 - 1	SASP2-123-1-10-37	81	0.0429	-	7471B	PPB	27 Dec 2021 11:28:37	U	SMPL
M5239-04 - 1	SASP2-123-1-40-57	771	0.1778	-	7471B	PPB	27 Dec 2021 11:30:52	U	SMPL
M5239-04DUP - 1	SASP2-123-1-40-57DUP	743	0.1723	-	7471B	PPB	27 Dec 2021 11:33:09	U	SMPL
M5239-04MS - 1	SASP2-123-1-40-57MS	21183	4.1681	-	7471B	PPB	27 Dec 2021 11:35:25	U	SMPL
M5239-04MSD - 1	SASP2-123-1-40-57MSD	20420	4.0190	-	7471B	PPB	27 Dec 2021 11:37:41	U	SMPL
M5218-01LX5 - 1		-5	0.0261	-	7471B	PPB	27 Dec 2021 11:39:58	U	SMPL
M5218-01A - 1		20499	4.0344	-	7471B	PPB	27 Dec 2021 11:42:15	U	SMPL
M5239-04LX5 - 1		166	0.0595	-	7471B	PPB	27 Dec 2021 11:44:33	U	SMPL
M5239-04A - 1		22688	4.4623	-	7471B	PPB	27 Dec 2021 11:46:48	U	SMPL
CCV99 - 1	CCV99	26308	5.1700	-	7471B	PPB	27 Dec 2021 11:49:04	U	SMPL
CCB99 - 1	CCB99	-63	0.0148	-	7471B	PPB	27 Dec 2021 11:51:20	U	SMPL