


LB 117978

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

41094



Linear

A= 0.0000e+000

B= 2.4273e-004 slope

C= -4.1245e-002 y-intercept

Rho= 0.9998340

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.006	-0.006	145	0.000	145					=
0.05	0.050					205	0				-5
0.20	0.200	0.191	-0.009	957	0.0 %	957					-3
2.50	2.500	2.429	-0.071	10177	0.0 %	10177					3
5.00	5.000	5.131	0.131	21307	0.0 %	21307					0
7.50	7.500	7.522	0.022	31159	0.0 %	31159					0
10.0	10.000	9.933	-0.067	41094	0.0 %	41094					-1

LB117978
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	145	-		0 SFAM01.1	PPB	12/27/2021 14:16	Std
	0.2 S01	957	-		0.2 SFAM01.1	PPB	12/27/2021 14:18	Std
	2.5 S02	10177	-		2.5 SFAM01.1	PPB	12/27/2021 14:20	Std
	5 S03	21307	-		5 SFAM01.1	PPB	12/27/2021 14:22	Std
	7.5 S04	31159	-		7.5 SFAM01.1	PPB	12/27/2021 14:25	Std
	10 S05	41094	-		10 SFAM01.1	PPB	12/27/2021 14:27	Std
ICV004	ICV004	17286	4.1546 -		SFAM01.1	PPB	12/27/2021 14:30	SMPL
ICB004	ICB004	96	-0.0179 -		SFAM01.1	PPB	12/27/2021 14:32	SMPL
CCV006	CCV006	21466	5.1692 -		SFAM01.1	PPB	12/27/2021 14:34	SMPL
CCB006	CCB006	103	-0.0162 -		SFAM01.1	PPB	12/27/2021 14:37	SMPL
PB141597BL	PBW597	-15	-0.0449 -		SFAM01.1	PPB	12/27/2021 14:39	SMPL
M5057-01	MH1F90	4	-0.0403 -		SFAM01.1	PPB	12/27/2021 14:41	SMPL
M5057-02	MH1F91	-6	-0.0427 -		SFAM01.1	PPB	12/27/2021 14:43	SMPL
M5057-03	MH1F92	19	-0.0366 -		SFAM01.1	PPB	12/27/2021 14:46	SMPL
M5057-04	MH1F93	18	-0.0369 -		SFAM01.1	PPB	12/27/2021 14:48	SMPL
M5057-05	MH1F94	11	-0.0386 -		SFAM01.1	PPB	12/27/2021 14:50	SMPL
M5057-06	MH1F95	26	-0.0349 -		SFAM01.1	PPB	12/27/2021 14:52	SMPL
M5057-07	MH1F96	122	-0.0116 -		SFAM01.1	PPB	12/27/2021 14:55	SMPL
M5057-08	MH1F97	111	-0.0143 -		SFAM01.1	PPB	12/27/2021 14:57	SMPL
M5057-09	MH1F98	30	-0.034 -		SFAM01.1	PPB	12/27/2021 14:59	SMPL
M5057-10	MH1F99	21	-0.0361 -		SFAM01.1	PPB	12/27/2021 15:01	SMPL
M5057-11	MH1G00	-3	-0.042 -		SFAM01.1	PPB	12/27/2021 15:04	SMPL
M5057-12	MH1G01	18	-0.0369 -		SFAM01.1	PPB	12/27/2021 15:06	SMPL
M5057-13	MH1G02	54	-0.0281 -		SFAM01.1	PPB	12/27/2021 15:08	SMPL
M5057-14	MH1G03	-5	-0.0425 -		SFAM01.1	PPB	12/27/2021 15:11	SMPL
M5057-15	MH1G04	23	-0.0357 -		SFAM01.1	PPB	12/27/2021 15:13	SMPL
M5057-16	MH1G05	52	-0.0286 -		SFAM01.1	PPB	12/27/2021 15:15	SMPL
M5057-17	MH1G06	2	-0.0408 -		SFAM01.1	PPB	12/27/2021 15:17	SMPL
M5057-18	MH1G07	27	-0.0347 -		SFAM01.1	PPB	12/27/2021 15:20	SMPL
M5057-19	MH1G07D	15	-0.0376 -		SFAM01.1	PPB	12/27/2021 15:22	SMPL
CCV007	CCV007	22271	5.3646 -		SFAM01.1	PPB	12/27/2021 15:24	SMPL
CCB007	CCB007	22	-0.0359 -		SFAM01.1	PPB	12/27/2021 15:26	SMPL
M5057-20	MH1G07S	4184	0.9743 -		SFAM01.1	PPB	12/27/2021 15:31	SMPL
PB141600BL	PBW600	1	-0.041 -		SFAM01.1	PPB	12/27/2021 15:33	SMPL
M5058-01	MH1G08	81	-0.0216 -		SFAM01.1	PPB	12/27/2021 15:36	SMPL
M5058-02	MH1G09	18	-0.0369 -		SFAM01.1	PPB	12/27/2021 15:38	SMPL
M5058-03	MH1G10	63	-0.026 -		SFAM01.1	PPB	12/27/2021 15:40	SMPL
M5058-04	MH1G11	11	-0.0386 -		SFAM01.1	PPB	12/27/2021 15:42	SMPL
M5058-05	MH1G12	24	-0.0354 -		SFAM01.1	PPB	12/27/2021 15:45	SMPL
M5058-06	MH1G13	46	-0.0301 -		SFAM01.1	PPB	12/27/2021 15:47	SMPL
M5058-07	MH1G14	56	-0.0277 -		SFAM01.1	PPB	12/27/2021 15:49	SMPL
M5058-08	MH1G15	38	-0.032 -		SFAM01.1	PPB	12/27/2021 15:52	SMPL
M5058-09	MH1G16	9	-0.0391 -		SFAM01.1	PPB	12/27/2021 15:54	SMPL
M5058-10	MH1G17	51	-0.0289 -		SFAM01.1	PPB	12/27/2021 15:56	SMPL
M5058-11	MH1G19	124	-0.0111 -		SFAM01.1	PPB	12/27/2021 15:58	SMPL
M5058-12	MH1G20	179	0.0022 -		SFAM01.1	PPB	12/27/2021 16:01	SMPL

LB117978
INSTRUMENT ID : CV1

M5058-13	MH1G21	356	0.0452 -	SFAM01.1	PPB	12/27/2021 16:03	SMPL
M5058-14	MH1G22	52	-0.0286 -	SFAM01.1	PPB	12/27/2021 16:05	SMPL
M5058-15	MH1G23	125	-0.0109 -	SFAM01.1	PPB	12/27/2021 16:07	SMPL
M5058-16	MH1G24	55	-0.0279 -	SFAM01.1	PPB	12/27/2021 16:10	SMPL
M5058-17	MH1G25	74	-0.0233 -	SFAM01.1	PPB	12/27/2021 16:12	SMPL
CCV008	CCV008	21237	5.1136 -	SFAM01.1	PPB	12/27/2021 16:16	SMPL
CCB008	CCB008	126	-0.0107 -	SFAM01.1	PPB	12/27/2021 16:19	SMPL
PB141601BL	PBW601	12	-0.0383 -	SFAM01.1	PPB	12/27/2021 16:24	SMPL
M5059-01	MH1G27	-12	-0.0442 -	SFAM01.1	PPB	12/27/2021 16:26	SMPL
M5059-02	MH1G28	5	-0.04 -	SFAM01.1	PPB	12/27/2021 16:28	SMPL
M5059-03	MH1G29	-8	-0.0432 -	SFAM01.1	PPB	12/27/2021 16:31	SMPL
M5059-04	MH1G30	-11	-0.0439 -	SFAM01.1	PPB	12/27/2021 16:33	SMPL
M5059-05	MH1G31	160	-0.0024 -	SFAM01.1	PPB	12/27/2021 16:35	SMPL
M5059-06	MH1G33	35	-0.0327 -	SFAM01.1	PPB	12/27/2021 16:37	SMPL
M5059-07	MH1G34	76	-0.0228 -	SFAM01.1	PPB	12/27/2021 16:40	SMPL
M5059-08	MH1G35	148	-0.0053 -	SFAM01.1	PPB	12/27/2021 16:42	SMPL
M5059-09	MH1G36	9	-0.0391 -	SFAM01.1	PPB	12/27/2021 16:44	SMPL
M5059-10	MH1G37	54	-0.0281 -	SFAM01.1	PPB	12/27/2021 16:46	SMPL
M5059-11	MH1G38	53	-0.0284 -	SFAM01.1	PPB	12/27/2021 16:49	SMPL
M5059-12	MH1G39	158	-0.0029 -	SFAM01.1	PPB	12/27/2021 16:51	SMPL
M5059-13	MH1G40	100	-0.017 -	SFAM01.1	PPB	12/27/2021 16:53	SMPL
M5059-14	MH1G41	94	-0.0184 -	SFAM01.1	PPB	12/27/2021 16:56	SMPL
M5059-15	MH1G42	100	-0.017 -	SFAM01.1	PPB	12/27/2021 16:58	SMPL
M5059-16	MH1G43	1419	0.3032 -	SFAM01.1	PPB	12/27/2021 17:00	SMPL
M5059-17	MH1G44	243	0.0177 -	SFAM01.1	PPB	12/27/2021 17:02	SMPL
CCV009	CCV009	21813	5.2534 -	SFAM01.1	PPB	12/27/2021 17:05	SMPL
CCB009	CCB009	123	-0.0114 -	SFAM01.1	PPB	12/27/2021 17:07	SMPL
M5059-18	MH1G45	42	-0.0311 -	SFAM01.1	PPB	12/27/2021 17:09	SMPL
M5059-19	MH1G45D	54	-0.0281 -	SFAM01.1	PPB	12/27/2021 17:11	SMPL
M5059-20	MH1G45S	3964	0.9209 -	SFAM01.1	PPB	12/27/2021 17:14	SMPL
PB141602BL	PBW602	71	-0.024 -	SFAM01.1	PPB	12/27/2021 17:16	SMPL
M5060-01	MH1G46	122	-0.0116 -	SFAM01.1	PPB	12/27/2021 17:18	SMPL
M5060-02	MH1G47	92	-0.0189 -	SFAM01.1	PPB	12/27/2021 17:21	SMPL
M5060-03	MH1G48	1260	0.2646 -	SFAM01.1	PPB	12/27/2021 17:23	SMPL
M5060-04	MH1G49	108	-0.015 -	SFAM01.1	PPB	12/27/2021 17:25	SMPL
M5060-05	MH1G50	695	0.1275 -	SFAM01.1	PPB	12/27/2021 17:27	SMPL
M5060-06	MH1G51	3827	0.8877 -	SFAM01.1	PPB	12/27/2021 17:30	SMPL
M5060-07	MH1G52	166	-0.001 -	SFAM01.1	PPB	12/27/2021 17:32	SMPL
M5060-08	MH1G53	135	-0.0085 -	SFAM01.1	PPB	12/27/2021 17:34	SMPL
M5060-09	MH1G54	108	-0.015 -	SFAM01.1	PPB	12/27/2021 17:36	SMPL
M5060-10	MH1G55	128	-0.0102 -	SFAM01.1	PPB	12/27/2021 17:39	SMPL
M5060-11	MH1G56	147	-0.0056 -	SFAM01.1	PPB	12/27/2021 17:41	SMPL
M5060-12	MH1G57	139	-0.0075 -	SFAM01.1	PPB	12/27/2021 17:43	SMPL
M5060-13	MH1G58	186	0.0039 -	SFAM01.1	PPB	12/27/2021 17:46	SMPL
M5060-14	MH1G59	491	0.0779 -	SFAM01.1	PPB	12/27/2021 17:48	SMPL
M5060-15	MH1G60	67	-0.025 -	SFAM01.1	PPB	12/27/2021 17:50	SMPL
M5060-16	MH1G61	161	-0.0022 -	SFAM01.1	PPB	12/27/2021 17:52	SMPL

LB117978

INSTRUMENT ID : CV1

CCV010	CCV010	22058	5.3129 -	SFAM01.1	PPB	12/27/2021 17:55 SMPL
CCB010	CCB010	139	-0.0075 -	SFAM01.1	PPB	12/27/2021 17:57 SMPL
M5060-17	MH1G61D	171	0.0003 -	SFAM01.1	PPB	12/27/2021 17:59 SMPL
M5060-18	MH1G61S	5572	1.3112 -	SFAM01.1	PPB	12/27/2021 18:01 SMPL
M5058-18	MH1G26	-18	-0.0456 -	SFAM01.1	PPB	12/27/2021 18:04 SMPL
M5058-19	MH1G26D	16	-0.0374 -	SFAM01.1	PPB	12/27/2021 18:06 SMPL
M5058-20	MH1G26S	4848	1.1355 -	SFAM01.1	PPB	12/27/2021 18:08 SMPL
CCV011	CCV011	21599	5.2015 -	SFAM01.1	PPB	12/27/2021 18:11 SMPL
CCB011	CCB011	117	-0.0128 -	SFAM01.1	PPB	12/27/2021 18:13 SMPL