



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117980

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/27/2021	14:30
2	CAL2	1	Water	NA	NA	20.7	7.00	12/27/2021	14:31
3	CAL3	1	Water	NA	NA	20.7	10.03	12/27/2021	14:32
4	ICV	1	Water	NA	NA	20.6	6.99	12/27/2021	14:35
5	CCV1	1	Water	NA	NA	20.6	2.01	12/27/2021	14:40
6	M5241-02	1	Solid	20.03	20	20.3	9.77	12/27/2021	14:45
7	M5241-05	1	Solid	20.02	20	20.2	7.97	12/27/2021	14:49
8	M5241-05DUP	1	Solid	20.03	20	20.3	7.98	12/27/2021	14:50
9	CCV2	1	Water	NA	NA	20.8	12.02	12/27/2021	15:00

12/11/2020

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m5241 WorkList ID : 155561 Department : Wet-Chemistry Date : 12-27-2021 14:20:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/27/2021	Solid	M5241-02	Corrosivity	Cool 4 deg C	PSEG03	I11	TP-2S	12/27/2021	9045D
12/27/2021	Solid	M5241-05	Corrosivity	Cool 4 deg C	PSEG03	I11	TP-2D	12/27/2021	9045D

Date/Time 12/11/2021 14:25
Raw Sample Received by: JP WC
Raw Sample Relinquished by: CP SM

Date/Time 12/11/21 14:00
Raw Sample Received by: JP WC
Raw Sample Relinquished by: JP WC