

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB117987

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP93829
COD calibration std. 100 ppm	WP93827
COD calibration std. 10 ppm	WP93825
COD calibration std. 150 ppm	WP93828
COD calibration std. 50 ppm	WP93826
COD calibration std. 0 ppm	WP93824
COD CCV std, 50ppm	WP93830
COD ICV-LCS std, 50ppm	WP94497
COD CCV std, 50ppm	WP94496
COD Digestion Vials Low Level 0-150Mg/L	W2827

Temp In (C): <u>150</u>	Date In: <u>12/28/2021</u>	Time In: <u>09:30</u>
Temp Out (C): <u>151</u>	Date Out: <u>12/28/2021</u>	Time Out: <u>11:30</u>

Intercept: -0.2481

Slope: 1.0105

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	0.246		11/09/2021	13:10
2	CAL2	10	1	Water	9.000	9.152	8.5	11/09/2021	13:10
3	CAL3	50	1	Water	51.000	50.716	-1.4	11/09/2021	13:11
4	CAL4	100	1	Water	101.000	100.196	-0.2	11/09/2021	13:11
5	CAL5	150	1	Water	151.000	149.676	0.2	11/09/2021	13:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB117987

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	50.000	49.726	12/28/2021	14:42
2	ICB		NA	NA	1	Water	0.000	0.246	12/28/2021	14:42
3	CCV1	50	NA	NA	1	Water	50.000	49.726	12/28/2021	14:43
4	CCB1		NA	NA	1	Water	0.000	0.246	12/28/2021	14:43
5	LB117987BL		NA	NA	1	Water	0.000	0.246	12/28/2021	14:44
6	LB117987BS	50	NA	NA	1	Water	50.000	49.726	12/28/2021	14:44
7	M5204-01		NA	NA	10	Water	29.000	28.944	12/28/2021	14:45
8	M5204-01DUP		NA	NA	10	Water	30.000	29.934	12/28/2021	14:45
9	M5204-01MS	50	NA	NA	10	Water	34.000	33.892	12/28/2021	14:46
10	M5204-01MSD	50	NA	NA	10	Water	35.000	34.882	12/28/2021	14:46
11	M5205-01		NA	NA	10	Water	60.000	59.622	12/28/2021	14:47
12	CCV2	50	NA	NA	1	Water	50.000	49.726	12/28/2021	14:47
13	CCB2		NA	NA	1	Water	0.000	0.246	12/28/2021	14:47