


LB117989

7470A

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.6151e-004

C= -2.1191e-002

Rho= 0.9998691

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.015	-0.015	23	0.000	23					-11
0.2	0.200	0.179	-0.021	766	0.0 %	766					1
2.5	2.500	2.528	0.028	9750	0.0 %	9750					0
5.0	5.000	4.976	-0.024	19111	0.0 %	19111					+2 2
7.5	7.500	7.613	0.113	29192	0.0 %	29192					-1
10.0	10.000	9.918	-0.082	38009	0.0 %	38009					nb

12128121

LB117989 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Dec 2021 11:19:56

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.00007470A	PPB		28 Dec 2021 11:24:14	S	Std
0.2 - 1	50.2	766	-	0.20007470A	PPB		28 Dec 2021 11:26:31	S	Std
2.5 - 1	52.5	9750	-	2.50007470A	PPB		28 Dec 2021 11:28:49	S	Std
5.0 - 1	55	19111	-	5.00007470A	PPB		28 Dec 2021 11:31:06	S	Std
7.5 - 1	57.5	29192	-	7.50007470A	PPB		28 Dec 2021 11:36:02	S	Std
10.0 - 1	50	38009	-	10.00007470A	PPB		28 Dec 2021 11:38:17	S	Std
ICV26 - 1	ICV26	15107	3.9294	-7470A	PPB		28 Dec 2021 11:40:58	U	SMPL
ICB26 - 1	ICB26	-65	-0.0382	-7470A	PPB		28 Dec 2021 11:43:12	U	SMPL
CCV12 - 1	CCV12	18860	4.9108	-7470A	PPB		28 Dec 2021 11:45:28	U	SMPL
CCB12 - 1	CCB12	-55	-0.0356	-7470A	PPB		28 Dec 2021 11:47:42	U	SMPL
CRA - 1	CRA	728	0.1692	-7470A	PPB		28 Dec 2021 11:49:58	U	SMPL
HighStd - 1	HighStd	37656	9.8261	-7470A	PPB		28 Dec 2021 11:52:13	U	SMPL
ChkStd - 1	ChkStd	26396	6.8815	-7470A	PPB		28 Dec 2021 11:54:28	U	SMPL
PB141712BL - 1	PBW	-67	-0.0387	-7470A	PPB		28 Dec 2021 11:56:44	U	SMPL
PB141712BS - 1	LCSW	14899	3.8750	-7470A	PPB		28 Dec 2021 11:59:00	U	SMPL
M5198-16 - 1	TP-1-5	-21	-0.0267	-7470A	PPB		28 Dec 2021 12:01:16	U	SMPL
M5230-02 - 1	PSC124042	14	-0.0175	-7470A	PPB		28 Dec 2021 12:03:33	U	SMPL
M5230-04 - 1	PSC124055	0	-0.0212	-7470A	PPB		28 Dec 2021 12:05:50	U	SMPL
M5230-06 - 1	PSC124030	40	-0.0107	-7470A	PPB		28 Dec 2021 12:08:06	U	SMPL
M5230-08 - 1	PSC164718	152	0.0186	-7470A	PPB		28 Dec 2021 12:10:22	U	SMPL
CCV13 - 1	CCV13	18604	4.8439	-7470A	PPB		28 Dec 2021 12:12:38	U	SMPL
CCB13 - 1	CCB13	-35	-0.0303	-7470A	PPB		28 Dec 2021 12:14:53	U	SMPL
M5230-10 - 1	PSC124019	16	-0.0170	-7470A	PPB		28 Dec 2021 12:17:09	U	SMPL
M5237-02 - 1	TP-1S	15	-0.0173	-7470A	PPB		28 Dec 2021 12:19:24	U	SMPL
M5237-05 - 1	TP-1D	-7	-0.0230	-7470A	PPB		28 Dec 2021 12:21:39	U	SMPL
M5237-05DUP - 1	TP-1DDUP	16	-0.0170	-7470A	PPB		28 Dec 2021 12:23:55	U	SMPL
M5237-05MS - 1	TP-1DMS	13536	3.5186	-7470A	PPB		28 Dec 2021 12:26:11	U	SMPL
M5237-05MSD - 1	TP-1DMSD	13640	3.5458	-7470A	PPB		28 Dec 2021 12:28:27	U	SMPL
PB141719BL - 1	PBW	-28	-0.0285	-7470A	PPB		28 Dec 2021 12:30:43	U	SMPL
PB141719BS - 1	LCSW	12731	3.3080	-7470A	PPB		28 Dec 2021 12:33:00	U	SMPL
M5234-01 - 1	P-1	-26	-0.0280	-7470A	PPB		28 Dec 2021 12:35:17	U	SMPL
M5234-01DUP - 1	P-1DUP	4	-0.0201	-7470A	PPB		28 Dec 2021 12:37:33	U	SMPL
CCV14 - 1	CCV14	18876	4.9150	-7470A	PPB		28 Dec 2021 12:39:49	U	SMPL
CCB14 - 1	CCB14	-47	-0.0335	-7470A	PPB		28 Dec 2021 12:42:04	U	SMPL
M5234-01MS - 1	P-1MS	12900	3.3522	-7470A	PPB		28 Dec 2021 12:44:20	U	SMPL
M5234-01MSD - 1	P-1MSD	13957	3.6287	-7470A	PPB		28 Dec 2021 12:46:35	U	SMPL
M5234-02 - 1	P-1-(FB)	-32	-0.0296	-7470A	PPB		28 Dec 2021 12:48:52	U	SMPL
PB141712TB - 1	PB141693TB	-4	-0.0222	-7470A	PPB		28 Dec 2021 12:51:07	U	SMPL
M5237-05LX5 - 1		2	-0.0207	-7470A	PPB		28 Dec 2021 12:53:23	U	SMPL
M5237-05A - 1		15380	4.0008	-7470A	PPB		28 Dec 2021 12:55:40	U	SMPL
M5234-01LX5 - 1		-42	-0.0322	-7470A	PPB		28 Dec 2021 12:57:57	U	SMPL
M5234-01A - 1		15513	4.0356	-7470A	PPB		28 Dec 2021 13:00:14	U	SMPL
CCV15 - 1	CCV15	18643	4.8541	-7470A	PPB		28 Dec 2021 13:02:31	U	SMPL
CCB15 - 1	CCB15	-36	-0.0306	-7470A	PPB		28 Dec 2021 13:04:47	U	SMPL