



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117991

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	12/21/2021	15:30
2	CAL2	1	Water	NA	NA	20.3	7.00	12/21/2021	15:31
3	CAL3	1	Water	NA	NA	20.3	10.03	12/21/2021	15:32
4	ICV	1	Water	NA	NA	20.4	7.02	12/21/2021	15:33
5	CCV1	1	Water	NA	NA	20.5	2.01	12/21/2021	15:35
6	M5197-05	1	Water	NA	NA	20.6	6.66	12/21/2021	15:36
7	M5197-05DUP	1	Water	NA	NA	20.7	6.67	12/21/2021	15:37
8	CCV2	1	Water	NA	NA	20.6	12.03	12/21/2021	15:38

VB 1179A1

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5197 W-1 WorkList ID : 155589 Department : Wet-Chemistry Date : 12-21-2021 15:10:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2021	Water	M5197-05	pH	Cool 4 deg C	PSEG03	H61	AUD-1233-1234	12/21/2021	9040C

Date/Time 12-21-21 15:15
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)

Date/Time 12-21-21 17:00
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)