

LB117992


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33894



A= 0.0000e+000

B= 2.9016e-004

C= -4.5853e-003

Rho= 0.9993510

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b100
0.0	0.000	0.001	0.001	20	0.000	20					-14
0.05	0.050	0.043	-0.007	165	0.0 %	165	0				-11
0.2	0.200	0.179	-0.021	634	0.0 %	634					2
2.5	2.500	2.541	0.041	8773	0.0 %	8773					-3
5.0	5.000	4.875	-0.125	16816	0.0 %	16816					4
7.5	7.500	7.780	0.280	26829	0.0 %	26829					-2
10.0	10.000	9.830	-0.170	33894	0.0 %	33894					

LB117992 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Dec 2021 14:03:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	20	-	0.0000	245.1	PPB	28 Dec 2021 14:15:52	S	Std
0.05 - 1	50.05	165	-	0.0500	245.1	PPB	28 Dec 2021 14:18:09	S	Std
0.2 - 1	50.2	634	-	0.2000	245.1	PPB	28 Dec 2021 14:20:26	S	Std
2.5 - 1	52.5	8773	-	2.5000	245.1	PPB	28 Dec 2021 14:22:43	S	Std
5.0 - 1	55	16816	-	5.0000	245.1	PPB	28 Dec 2021 14:30:30	S	Std
7.5 - 1	57.5	26829	-	7.5000	245.1	PPB	28 Dec 2021 14:32:46	S	Std
10.0 - 1	510	33894	-	10.0000	245.1	PPB	28 Dec 2021 14:37:35	S	Std
ICV27 - 1	ICV27	13790	3.9967	-245.1		PPB	28 Dec 2021 14:40:20	U	SMPL
ICB27 - 1	ICB27	-27	-0.0124	-245.1		PPB	28 Dec 2021 14:42:35	U	SMPL
CCV16 - 1	CCV16	16733	4.8507	-245.1		PPB	28 Dec 2021 14:44:50	U	SMPL
CCB16 - 1	CCB16	-32	-0.0139	-245.1		PPB	28 Dec 2021 14:47:06	U	SMPL
CRA - 1	CRA	145	0.0375	-245.1		PPB	28 Dec 2021 14:49:21	U	SMPL
HighStd - 1	HighStd	33537	9.7266	-245.1		PPB	28 Dec 2021 14:51:36	U	SMPL
ChkStd - 1	ChkStd	25658	7.4404	-245.1		PPB	28 Dec 2021 14:57:50	U	SMPL
PB141742BL - 1	PBW	-57	-0.0211	-245.1		PPB	28 Dec 2021 15:00:05	U	SMPL
PB141742BS - 1	LCSW	14507	4.2048	-245.1		PPB	28 Dec 2021 15:02:21	U	SMPL
M5176-01 - 1	TI-658-EB	-44	-0.0174	-245.1		PPB	28 Dec 2021 15:04:37	U	SMPL
M5176-03 - 1	TI-658	63	0.0137	-245.1		PPB	28 Dec 2021 15:06:54	U	SMPL
M5176-04 - 1	TI-658-X	67	0.0149	-245.1		PPB	28 Dec 2021 15:09:11	U	SMPL
M5176-04DUP - 1	TI-658-XDUP	83	0.0195	-245.1		PPB	28 Dec 2021 15:11:28	U	SMPL
M5176-04MS - 1	TI-658-XMS	15596	4.5208	-245.1		PPB	28 Dec 2021 15:13:43	U	SMPL
CCV17 - 1	CCV17	16914	4.9032	-245.1		PPB	28 Dec 2021 15:15:58	U	SMPL
CCB17 - 1	CCB17	-37	-0.0153	-245.1		PPB	28 Dec 2021 15:18:13	U	SMPL
M5176-04MSD - 1	TI-658-XMSD	14214	4.1198	-245.1		PPB	28 Dec 2021 15:20:28	U	SMPL
M5176-04LX5 - 1		-9	-0.0072	-245.1		PPB	28 Dec 2021 15:22:43	U	SMPL
M5176-04A - 1		13887	4.0249	-245.1		PPB	28 Dec 2021 15:24:59	U	SMPL
CCV18 - 1	CCV18	16835	4.8803	-245.1		PPB	28 Dec 2021 15:27:15	U	SMPL
CCB18 - 1	CCB18	-31	-0.0136	-245.1		PPB	28 Dec 2021 15:29:31	U	SMPL