

LB118003


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

37748



A= 0.0000e+000

B= 2.6519e-004

C= 5.4094e-002

Rho= 0.9996788

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.061	0.061	26	0.000	26					12
0.2	0.200	0.224	0.024	640	0.0 %	640					1
2.5	2.500	2.519	0.019	9296	0.0 %	9296					-4
5.0	5.000	4.794	-0.206	17875	0.0 %	17875					0
7.5	7.500	7.537	0.037	28217	0.0 %	28217					1
10.0	10.000	10.065	0.065	37748	0.0 %	37748					

LB118003 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Dec 2021 11:18:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	29 Dec 2021 11:30:28	S	Std
0.2 - 1	50.2	640	-	0.2000	7470A	PPB	29 Dec 2021 11:32:44	S	Std
2.5 - 1	52.5	9296	-	2.5000	7470A	PPB	29 Dec 2021 11:35:01	S	Std
5.0 - 1	55	17875	-	5.0000	7470A	PPB	29 Dec 2021 11:39:51	S	Std
7.5 - 1	57.5	28217	-	7.5000	7470A	PPB	29 Dec 2021 11:42:07	S	Std
10.0 - 1	510	37748	-	10.0000	7470A	PPB	29 Dec 2021 11:44:23	S	Std
ICV28 - 1	ICV28	14886	4.0017	-	7470A	PPB	29 Dec 2021 11:47:25	U	SMPL
ICB28 - 1	ICB28	-48	0.0414	-	7470A	PPB	29 Dec 2021 11:49:40	U	SMPL
CCV19 - 1	CCV19	17679	4.7424	-	7470A	PPB	29 Dec 2021 11:51:55	U	SMPL
CCB19 - 1	CCB19	0	0.0541	-	7470A	PPB	29 Dec 2021 11:54:10	U	SMPL
CRA - 1	CRA	633	0.2220	-	7470A	PPB	29 Dec 2021 11:56:25	U	SMPL
HighStd - 1	HighStd	38137	10.1677	-	7470A	PPB	29 Dec 2021 11:58:40	U	SMPL
ChkStd - 1	ChkStd	26080	6.9703	-	7470A	PPB	29 Dec 2021 12:00:55	U	SMPL
PB141741BL - 1	PBW	-74	0.0345	-	7470A	PPB	29 Dec 2021 12:03:11	U	SMPL
PB141741BS - 1	LCSW	15285	4.1075	-	7470A	PPB	29 Dec 2021 12:05:27	U	SMPL
M5241-02 - 1	TP-2S	-6	0.0525	-	7470A	PPB	29 Dec 2021 12:07:44	U	SMPL
M5241-05 - 1	TP-2D	8	0.0562	-	7470A	PPB	29 Dec 2021 12:10:00	U	SMPL
M5241-05DUP - 1	TP-2DDUP	21	0.0597	-	7470A	PPB	29 Dec 2021 12:12:17	U	SMPL
M5241-05MS - 1	TP-2DMS	15051	4.0455	-	7470A	PPB	29 Dec 2021 12:14:34	U	SMPL
M5241-05MSD - 1	TP-2DMSD	13395	3.6063	-	7470A	PPB	29 Dec 2021 12:16:49	U	SMPL
CCV20 - 1	CCV20	17409	4.6708	-	7470A	PPB	29 Dec 2021 12:19:04	U	SMPL
CCB20 - 1	CCB20	-54	0.0398	-	7470A	PPB	29 Dec 2021 12:21:19	U	SMPL
PB141741TB - 1	PB141720TB	2	0.0546	-	7470A	PPB	29 Dec 2021 12:23:35	U	SMPL
M5241-05LX5 - 1		2	0.0546	-	7470A	PPB	29 Dec 2021 12:25:50	U	SMPL
M5241-05A - 1		15708	4.2197	-	7470A	PPB	29 Dec 2021 12:28:05	U	SMPL
CCV21 - 1	CCV21	17539	4.7053	-	7470A	PPB	29 Dec 2021 12:30:21	U	SMPL
CCB21 - 1	CCB21	-30	0.0461	-	7470A	PPB	29 Dec 2021 12:32:37	U	SMPL