

L3118008


SFAM01.1

INSTRUMENT ID: CVZ

Linear

μ Abs.:

44160



A= 0.0000e+000

B= 2.2745e-004 *slope*

C= -7.2693e-002 *y-intercept*

Rho= 0.9997654

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>9/6/1</i>
0.00	0.000	-0.049	-0.049	104	0.000	104					-
0.05	0.050					205	0				16
0.20	0.200	0.231	0.031	1334	0.0 %	1334					4
2.50	2.500	2.589	0.089	11702	0.0 %	11702					3
5.00	5.000	4.866	-0.134	21715	0.0 %	21715					1
7.50	7.500	7.592	0.092	33697	0.0 %	33697					0
10.0	10.000	9.971	-0.029	44160	0.0 %	44160					

LB118008
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	104	-		0 SFAM01.1	PPB	12/29/2021 13:28	Std
	0.2 S01	1334	-		0.2 SFAM01.1	PPB	12/29/2021 13:30	Std
	2.5 S02	11702	-		2.5 SFAM01.1	PPB	12/29/2021 13:35	Std
	5 S03	21715	-		5 SFAM01.1	PPB	12/29/2021 13:43	Std
	7.5 S04	33697	-		7.5 SFAM01.1	PPB	12/29/2021 13:45	Std
	10 S05	44160	-		10 SFAM01.1	PPB	12/29/2021 13:50	Std
ICV005	ICV005	18433	4.1199 -		SFAM01.1	PPB	12/29/2021 13:53	SMPL
ICB005	ICB005	39	-0.0638 -		SFAM01.1	PPB	12/29/2021 13:55	SMPL
CCV021	CCV021	23322	5.2319 -		SFAM01.1	PPB	12/29/2021 13:57	SMPL
CCB021	CCB021	23	-0.0675 -		SFAM01.1	PPB	12/29/2021 13:59	SMPL
PB141755BL	PBS755	38	-0.0641 -		SFAM01.1	PPB	12/29/2021 14:02	SMPL
M5225-01	GBC27	96	-0.0509 -		SFAM01.1	PPB	12/29/2021 14:04	SMPL
M5225-02	GBC27D	66	-0.0577 -		SFAM01.1	PPB	12/29/2021 14:06	SMPL
M5225-03	GBC27S	10951	2.4181 -		SFAM01.1	PPB	12/29/2021 14:08	SMPL
M5225-04	GBC28	467	0.0335 -		SFAM01.1	PPB	12/29/2021 14:11	SMPL
M5225-05	GBC29	474	0.0351 -		SFAM01.1	PPB	12/29/2021 14:13	SMPL
M5225-06	GBC30	2631	0.5257 -		SFAM01.1	PPB	12/29/2021 14:15	SMPL
M5225-07	GBC31	1831	0.3438 -		SFAM01.1	PPB	12/29/2021 14:17	SMPL
M5225-08	GBC32	1085	0.1741 -		SFAM01.1	PPB	12/29/2021 14:20	SMPL
M5225-09	GBC33	1757	0.3269 -		SFAM01.1	PPB	12/29/2021 14:22	SMPL
M5225-10	GBC34	1228	0.2066 -		SFAM01.1	PPB	12/29/2021 14:24	SMPL
M5225-11	GBC35	1172	0.1939 -		SFAM01.1	PPB	12/29/2021 14:27	SMPL
M5225-12	GBC36	501	0.0413 -		SFAM01.1	PPB	12/29/2021 14:29	SMPL
M5225-13	GBC37	480	0.0365 -		SFAM01.1	PPB	12/29/2021 14:31	SMPL
M5225-14	GBC38	626	0.0697 -		SFAM01.1	PPB	12/29/2021 14:33	SMPL
M5225-15	GBC39	467	0.0335 -		SFAM01.1	PPB	12/29/2021 14:36	SMPL
M5225-16	GBC40	566	0.056 -		SFAM01.1	PPB	12/29/2021 14:38	SMPL
M5225-17	GBC41	350	0.0069 -		SFAM01.1	PPB	12/29/2021 14:40	SMPL
M5225-18	GBC42	498	0.0406 -		SFAM01.1	PPB	12/29/2021 14:42	SMPL
M5225-19	GBC43	797	0.1086 -		SFAM01.1	PPB	12/29/2021 14:45	SMPL
CCV022	CCV022	23920	5.3679 -		SFAM01.1	PPB	12/29/2021 14:47	SMPL
CCB022	CCB022	24	-0.0672 -		SFAM01.1	PPB	12/29/2021 14:49	SMPL
M5225-20	GBC44	1099	0.1773 -		SFAM01.1	PPB	12/29/2021 14:51	SMPL
M5225-21	GBC45	1015	0.1582 -		SFAM01.1	PPB	12/29/2021 14:54	SMPL
PB141756BL	PBS756	41	-0.0634 -		SFAM01.1	PPB	12/29/2021 14:56	SMPL
M5226-01	GBC46	1210	0.2025 -		SFAM01.1	PPB	12/29/2021 14:58	SMPL
M5226-02	GBC47	659	0.0772 -		SFAM01.1	PPB	12/29/2021 15:01	SMPL
M5226-03	GBC48	1142	0.1871 -		SFAM01.1	PPB	12/29/2021 15:03	SMPL
M5226-04	GBC48D	1200	0.2002 -		SFAM01.1	PPB	12/29/2021 15:05	SMPL
M5226-05	GBC48S	12954	2.8737 -		SFAM01.1	PPB	12/29/2021 15:07	SMPL
M5226-06	GBC49	904	0.1329 -		SFAM01.1	PPB	12/29/2021 15:10	SMPL
M5226-07	GBC50	1327	0.2291 -		SFAM01.1	PPB	12/29/2021 15:12	SMPL
M5226-08	GBC51	3549	0.7345 -		SFAM01.1	PPB	12/29/2021 15:14	SMPL
M5226-09	GBC52	1031	0.1618 -		SFAM01.1	PPB	12/29/2021 15:17	SMPL
M5226-10	GBC53	964	0.1466 -		SFAM01.1	PPB	12/29/2021 15:19	SMPL
M5226-11	GBC54	966	0.147 -		SFAM01.1	PPB	12/29/2021 15:21	SMPL

LB118008
INSTRUMENT ID : CV1

M5226-12	GBC55	5796	1.2456 -	SFAM01.1	PPB	12/29/2021 15:23 SMPL
M5226-13	GBC56	1490	0.2662 -	SFAM01.1	PPB	12/29/2021 15:26 SMPL
M5226-14	GBC57	1781	0.3324 -	SFAM01.1	PPB	12/29/2021 15:28 SMPL
M5226-15	GBC58	1011	0.1573 -	SFAM01.1	PPB	12/29/2021 15:30 SMPL
M5226-16	GBC59	1054	0.167 -	SFAM01.1	PPB	12/29/2021 15:32 SMPL
M5226-17	GBC60	737	0.0949 -	SFAM01.1	PPB	12/29/2021 15:35 SMPL
CCV023	CCV023	23838	5.3492 -	SFAM01.1	PPB	12/29/2021 15:37 SMPL
CCB023	CCB023	33	-0.0652 -	SFAM01.1	PPB	12/29/2021 15:39 SMPL
M5226-18	GBC61	375	0.0126 -	SFAM01.1	PPB	12/29/2021 15:41 SMPL
M5226-19	GBC62	654	0.0761 -	SFAM01.1	PPB	12/29/2021 15:44 SMPL
M5226-20	GBC63	1044	0.1648 -	SFAM01.1	PPB	12/29/2021 15:46 SMPL
M5226-21	GBC64	2016	0.3858 -	SFAM01.1	PPB	12/29/2021 15:48 SMPL
CCV024	CCV024	24003	5.3867 -	SFAM01.1	PPB	12/29/2021 15:51 SMPL
CCB024	CCB024	38	-0.0641 -	SFAM01.1	PPB	12/29/2021 15:53 SMPL