



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118016

Slope : 99.2

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	12/29/2021	17:10
2	CAL2	1	Water	NA	NA	20.6	7.00	12/29/2021	17:11
3	CAL3	1	Water	NA	NA	20.6	10.03	12/29/2021	17:12
4	ICV	1	Water	NA	NA	20.7	6.99	12/29/2021	17:15
5	CCV1	1	Water	NA	NA	20.6	2.01	12/29/2021	17:20
6	M5254-02	1	Solid	20.03	20	22.5	8.32	12/29/2021	17:26
7	M5254-02DUP	1	Solid	20.02	20	22.6	8.33	12/29/2021	17:27
8	M5257-02	1	Solid	20.01	20	21.3	8.42	12/29/2021	17:28
9	CCV2	1	Water	NA	NA	20.3	12.03	12/29/2021	17:29

VP118016

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity m5257 WorkList ID : 155622 Department : Wet-Chemistry Date : 12-29-2021 16:18:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/29/2021	Solid	M5254-02	Corrosivity	Cool 4 deg C	PSEG03	H11	361	12/29/2021	9045D
12/29/2021	Solid	M5257-02	Corrosivity	Cool 4 deg C	ENTA05	D11	D3-12292021	12/29/2021	9045D

Date/Time 12/29/21 16:20
 Raw Sample Received by: JP/WC
 Raw Sample Relinquished by: CP SM

Date/Time 12/29/21 17:30
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP/WC