


LB118019

SFAM01.1

INSTRUMENT ID: EV1



A= 0.0000e+000

B= 2.3287e-004 slope

C= -8.8950e-002 y-intercept

Rho= 0.9996888

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.046	-0.046	183	0.000	183					5
0.05	0.050					205	0				3 4
0.20	0.200	0.207	0.007	1271	0.0 %	1271					4
2.50	2.500	2.592	0.092	11514	0.0 %	11514					1/11/22
5.00	5.000	4.880	-0.120	21336	0.0 %	21336					-2
7.50	7.500	7.642	0.142	33197	0.0 %	33197					2
10.0	10.000	9.926	-0.074	43004	0.0 %	43004					-1

LB118019
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	183	-		0 SFAM01.1	PPB	12/30/2021 10:22	Std
	0.2 S01	1271	-		0.2 SFAM01.1	PPB	12/30/2021 10:24	Std
	2.5 S02	11514	-		2.5 SFAM01.1	PPB	12/30/2021 10:27	Std
	5 S03	21336	-		5 SFAM01.1	PPB	12/30/2021 10:31	Std
	7.5 S04	33197	-		7.5 SFAM01.1	PPB	12/30/2021 10:34	Std
	10 S05	43004	-		10 SFAM01.1	PPB	12/30/2021 10:36	Std
ICV006	ICV006	17888	4.0767 -		SFAM01.1	PPB	12/30/2021 10:39	SMPL
ICB006	ICB006	177	-0.0477 -		SFAM01.1	PPB	12/30/2021 10:41	SMPL
CCV025	CCV025	21382	4.8904 -		SFAM01.1	PPB	12/30/2021 10:43	SMPL
CCB025	CCB025	168	-0.0498 -		SFAM01.1	PPB	12/30/2021 10:46	SMPL
PB141771BL	PBS771	79	-0.0706 -		SFAM01.1	PPB	12/30/2021 10:48	SMPL
M5227-01	GBC65	1368	0.2296 -		SFAM01.1	PPB	12/30/2021 10:50	SMPL
M5227-02	GBC66	556	0.0405 -		SFAM01.1	PPB	12/30/2021 10:52	SMPL
M5227-03	GBC67	295	-0.0203 -		SFAM01.1	PPB	12/30/2021 10:55	SMPL
M5227-04	GBC68	1168	0.183 -		SFAM01.1	PPB	12/30/2021 10:57	SMPL
M5227-05	GBC68D	1167	0.1828 -		SFAM01.1	PPB	12/30/2021 10:59	SMPL
M5227-06	GBC68S	12867	2.9074 -		SFAM01.1	PPB	12/30/2021 11:02	SMPL
M5227-07	GBC69	887	0.1176 -		SFAM01.1	PPB	12/30/2021 11:04	SMPL
M5227-08	GBC70	376	-0.0014 -		SFAM01.1	PPB	12/30/2021 11:06	SMPL
M5227-09	GBC71	2107	0.4017 -		SFAM01.1	PPB	12/30/2021 11:08	SMPL
M5227-10	GBC72	3300	0.6795 -		SFAM01.1	PPB	12/30/2021 11:11	SMPL
M5227-11	GBC73	526	0.0335 -		SFAM01.1	PPB	12/30/2021 11:13	SMPL
M5227-12	GBC74	2754	0.5524 -		SFAM01.1	PPB	12/30/2021 11:15	SMPL
M5227-13	GBC75	1022	0.149 -		SFAM01.1	PPB	12/30/2021 11:17	SMPL
M5227-14	GBC76	488	0.0247 -		SFAM01.1	PPB	12/30/2021 11:20	SMPL
M5227-15	GBC77	676	0.0685 -		SFAM01.1	PPB	12/30/2021 11:22	SMPL
M5227-16	GBC78	550	0.0391 -		SFAM01.1	PPB	12/30/2021 11:24	SMPL
M5227-17	GBC79	501	0.0277 -		SFAM01.1	PPB	12/30/2021 11:26	SMPL
M5227-18	GBC80	538	0.0363 -		SFAM01.1	PPB	12/30/2021 11:29	SMPL
M5227-19	GBC81	1277	0.2084 -		SFAM01.1	PPB	12/30/2021 11:31	SMPL
CCV026	CCV026	21596	4.9402 -		SFAM01.1	PPB	12/30/2021 11:33	SMPL
CCB026	CCB026	152	-0.0536 -		SFAM01.1	PPB	12/30/2021 11:36	SMPL
M5227-20	GBC82	725	0.0799 -		SFAM01.1	PPB	12/30/2021 11:38	SMPL
M5227-21	GBC83	615	0.0543 -		SFAM01.1	PPB	12/30/2021 11:40	SMPL
PB141772BL	PBS772	71	-0.0724 -		SFAM01.1	PPB	12/30/2021 11:42	SMPL
M5228-01	GBC84	175	-0.0482 -		SFAM01.1	PPB	12/30/2021 11:45	SMPL
M5228-02	GBC85	142	-0.0559 -		SFAM01.1	PPB	12/30/2021 11:47	SMPL
M5228-03	GBC86	423	0.0096 -		SFAM01.1	PPB	12/30/2021 11:49	SMPL
M5228-04	GBC87	332	-0.0116 -		SFAM01.1	PPB	12/30/2021 11:51	SMPL
M5228-05	GBC88	462	0.0186 -		SFAM01.1	PPB	12/30/2021 11:54	SMPL
M5228-06	GBC88D	454	0.0168 -		SFAM01.1	PPB	12/30/2021 11:56	SMPL
M5228-07	GBC88S	11853	2.6713 -		SFAM01.1	PPB	12/30/2021 11:58	SMPL
M5228-08	GBC89	843	0.1074 -		SFAM01.1	PPB	12/30/2021 12:00	SMPL
M5228-09	GBC90	445	0.0147 -		SFAM01.1	PPB	12/30/2021 12:03	SMPL
M5228-10	GBC91	141	-0.0561 -		SFAM01.1	PPB	12/30/2021 12:05	SMPL
M5228-11	GBC92	327	-0.0128 -		SFAM01.1	PPB	12/30/2021 12:07	SMPL

LB118019
INSTRUMENT ID : CV1

M5228-12	GBC93	438	0.013 -	SFAM01.1	PPB	12/30/2021 12:10	SMPL
M5228-13	GBC94	511	0.03 -	SFAM01.1	PPB	12/30/2021 12:12	SMPL
M5228-14	GBC95	1334	0.2217 -	SFAM01.1	PPB	12/30/2021 12:14	SMPL
M5228-15	GBC96	817	0.1013 -	SFAM01.1	PPB	12/30/2021 12:16	SMPL
M5228-16	GBC97	804	0.0983 -	SFAM01.1	PPB	12/30/2021 12:19	SMPL
M5228-17	GBC98	235	-0.0342 -	SFAM01.1	PPB	12/30/2021 12:21	SMPL
CCV027	CCV027	21799	4.9875 -	SFAM01.1	PPB	12/30/2021 12:23	SMPL
CCB027	CCB027	162	-0.0512 -	SFAM01.1	PPB	12/30/2021 12:25	SMPL
M5228-18	GBC99	542	0.0373 -	SFAM01.1	PPB	12/30/2021 12:28	SMPL
M5228-19	GBCA0	624	0.0564 -	SFAM01.1	PPB	12/30/2021 12:30	SMPL
M5228-20	GBCA1	685	0.0706 -	SFAM01.1	PPB	12/30/2021 12:32	SMPL
M5228-21	GBCA2	1772	0.3237 -	SFAM01.1	PPB	12/30/2021 12:34	SMPL
PB141773BL	PBS773	68	-0.0731 -	SFAM01.1	PPB	12/30/2021 12:37	SMPL
M5229-01	GBCA3	1146	0.1779 -	SFAM01.1	PPB	12/30/2021 12:39	SMPL
M5229-02	GBCA4	597	0.0501 -	SFAM01.1	PPB	12/30/2021 12:41	SMPL
M5229-03	GBCA5	1051	0.1558 -	SFAM01.1	PPB	12/30/2021 12:44	SMPL
M5229-04	GBCA6	1057	0.1572 -	SFAM01.1	PPB	12/30/2021 12:46	SMPL
M5229-05	GBCA7	632	0.0582 -	SFAM01.1	PPB	12/30/2021 12:48	SMPL
M5229-06	GBCA8	1016	0.1476 -	SFAM01.1	PPB	12/30/2021 12:50	SMPL
M5229-07	GBCA8D	1087	0.1642 -	SFAM01.1	PPB	12/30/2021 12:53	SMPL
M5229-08	GBCA8S	14204	3.2188 -	SFAM01.1	PPB	12/30/2021 12:55	SMPL
M5229-09	GBCA9	773	0.0911 -	SFAM01.1	PPB	12/30/2021 12:57	SMPL
M5229-10	GBCB0	866	0.1127 -	SFAM01.1	PPB	12/30/2021 12:59	SMPL
M5229-11	GBCB1	385	0.0007 -	SFAM01.1	PPB	12/30/2021 13:02	SMPL
M5229-12	GBCB2	561	0.0417 -	SFAM01.1	PPB	12/30/2021 13:04	SMPL
M5229-13	GBCB3	327	-0.0128 -	SFAM01.1	PPB	12/30/2021 13:06	SMPL
M5229-14	GBCB4	725	0.0799 -	SFAM01.1	PPB	12/30/2021 13:09	SMPL
M5229-15	GBCB5	159	-0.0519 -	SFAM01.1	PPB	12/30/2021 13:11	SMPL
CCV028	CCV028	21886	5.0077 -	SFAM01.1	PPB	12/30/2021 13:13	SMPL
CCB028	CCB028	168	-0.0498 -	SFAM01.1	PPB	12/30/2021 13:15	SMPL
M5229-16	GBCB6	627	0.0571 -	SFAM01.1	PPB	12/30/2021 13:18	SMPL
M5229-17	GBCB7	401	0.0044 -	SFAM01.1	PPB	12/30/2021 13:20	SMPL
M5229-18	GBCB8	681	0.0696 -	SFAM01.1	PPB	12/30/2021 13:22	SMPL
M5229-19	GBCB9	289	-0.0216 -	SFAM01.1	PPB	12/30/2021 13:24	SMPL
CCV029	CCV029	21754	4.977 -	SFAM01.1	PPB	12/30/2021 13:27	SMPL
CCB029	CCB029	167	-0.0501 -	SFAM01.1	PPB	12/30/2021 13:29	SMPL