



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118020

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/30/2021	10:26
2	CAL2	1	Water	NA	NA	20.8	7.01	12/30/2021	10:26
3	CAL3	1	Water	NA	NA	20.8	10.03	12/30/2021	10:27
4	ICV	1	Water	NA	NA	20.7	7.01	12/30/2021	10:29
5	CCV1	1	Water	NA	NA	20.7	2.01	12/30/2021	10:30
6	M5256-01	1	Solid	20.03	20	21.9	7.50	12/30/2021	10:33
7	M5256-01DUP	1	Solid	20.04	20	21.8	7.51	12/30/2021	10:34
8	CCV2	1	Water	NA	NA	20.6	12.03	12/30/2021	10:35

118020

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M5256 S WorkList ID : 155643 Department : Wet-Chemistry Date : 12-30-2021 10:17:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/29/2021	Solid	M5256-01	F pH	Cool 4 deg C	PSEG05	I11	BU-03-122921	12/29/2021	9045D

Date/Time 12/30/21 10:20
 Raw Sample Received by: RJ C E X T (266)
 Raw Sample Relinquished by: RJ C E X T (266)

Date/Time 12/30/21 11:30
 Raw Sample Received by: RJ C E X T (266)
 Raw Sample Relinquished by: RJ C E X T (266)