

LB118030


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

28176



A= 0.0000e+000

B= 3.5543e-004

C= -2.0116e-002

Rho= 0.9997634

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.007	-0.007	36	0.000	36					-
0.2	0.200	0.212	0.012	652	0.0 %	652					6
2.5	2.500	2.579	0.079	7312	0.0 %	7312					3
5.0	5.000	4.840	-0.160	13675	0.0 %	13675					-3
7.5	7.500	7.582	0.082	21389	0.0 %	21389					1
10.0	10.000	9.994	-0.006	28176	0.0 %	28176					0

LB118030 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jan 2022 09:51:06

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	36	-	0.0000	7470A	PPB	03 Jan 2022 10:30:10	S	Std
0.2 - 1	50.2	652	-	0.2000	7470A	PPB	03 Jan 2022 10:35:03	S	Std
2.5 - 1	52.5	7312	-	2.5000	7470A	PPB	03 Jan 2022 10:37:20	S	Std
5.0 - 1	15	13675	-	5.0000	7470A	PPB	03 Jan 2022 10:39:39	S	Std
7.5 - 1	57.5	21389	-	7.5000	7470A	PPB	03 Jan 2022 10:47:05	S	Std
10.0 - 1	520	28176	-	10.0000	7470A	PPB	03 Jan 2022 10:54:31	S	Std
ICV30 - 1	ICV30	12154	4.2998	-7470A	PPB	PPB	03 Jan 2022 10:59:51	U	SMPL
ICB30 - 1	ICB30	-13	-0.0247	-7470A	PPB	PPB	03 Jan 2022 11:02:05	U	SMPL
CCV34 - 1	CCV34	14126	5.0007	-7470A	PPB	PPB	03 Jan 2022 11:04:20	U	SMPL
CCB34 - 1	CCB34	-109	-0.0589	-7470A	PPB	PPB	03 Jan 2022 11:06:35	U	SMPL
CRA - 1	CRA	660	0.2145	-7470A	PPB	PPB	03 Jan 2022 11:08:51	U	SMPL
HighStd - 1	HighStd	29576	10.4920	-7470A	PPB	PPB	03 Jan 2022 11:11:07	U	SMPL
ChkStd - 1	ChkStd	20887	7.4037	-7470A	PPB	PPB	03 Jan 2022 11:13:22	U	SMPL
PB141776BL - 1	PBW	-146	-0.0720	-7470A	PPB	PPB	03 Jan 2022 11:15:40	U	SMPL
PB141776BS - 1	LCSW	11265	3.9838	-7470A	PPB	PPB	03 Jan 2022 11:20:28	U	SMPL
M5254-02 - 1	361	12	-0.0159	-7470A	PPB	PPB	03 Jan 2022 11:22:44	U	SMPL
M5257-02 - 1	D3-12292021	103	0.0165	-7470A	PPB	PPB	03 Jan 2022 11:25:00	U	SMPL
M5257-03 - 1	P2-A	3969	1.3906	-7470A	PPB	PPB	03 Jan 2022 11:30:34	U	SMPL
M5257-04 - 1	P2-B	342	0.1014	-7470A	PPB	PPB	03 Jan 2022 11:32:50	U	SMPL
M5257-05 - 1	P2-C	1082	0.3645	-7470A	PPB	PPB	03 Jan 2022 11:35:06	U	SMPL
CCV35 - 1	CCV35	15027	5.3209	-7470A	PPB	PPB	03 Jan 2022 11:37:23	U	SMPL
CCB35 - 1	CCB35	-75	-0.0468	-7470A	PPB	PPB	03 Jan 2022 11:39:41	U	SMPL
M5257-06 - 1	P2-D	241879	85.9505	-7470A	PPB	PPB	03 Jan 2022 11:41:56	U	SMPL
M5257-07 - 1	P3-A	2724	0.9481	-7470A	PPB	PPB	03 Jan 2022 11:44:11	U	SMPL
M5257-08 - 1	P3-B	248	0.0680	-7470A	PPB	PPB	03 Jan 2022 11:46:31	U	SMPL
M5257-08DUP - 1	P3-BDUP	243	0.0663	-7470A	PPB	PPB	03 Jan 2022 11:48:47	U	SMPL
M5257-08MS - 1	P3-BMS	12521	4.4302	-7470A	PPB	PPB	03 Jan 2022 11:53:33	U	SMPL
M5257-08MSD - 1	P3-BMSD	11825	4.1828	-7470A	PPB	PPB	03 Jan 2022 11:55:49	U	SMPL
PB141177TB - 1	PB141757TB	2	-0.0194	-7470A	PPB	PPB	03 Jan 2022 11:58:05	U	SMPL
M5257-08LX5 - 1		67	0.0037	-7470A	PPB	PPB	03 Jan 2022 12:00:22	U	SMPL
M5257-08A - 1		12304	4.3531	-7470A	PPB	PPB	03 Jan 2022 12:02:40	U	SMPL
M5257-06DLX10 - 1		26196	9.2907	-7470A	PPB	PPB	03 Jan 2022 12:07:42	U	SMPL
CCV36 - 1	CCV36	14794	5.2381	-7470A	PPB	PPB	03 Jan 2022 12:12:28	U	SMPL
CCB36 - 1	CCB36	-73	-0.0461	-7470A	PPB	PPB	03 Jan 2022 12:14:43	U	SMPL