



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118031

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	12/30/2021	18:00
2	CAL2	1	Water	NA	NA	20.9	7.00	12/30/2021	18:01
3	CAL3	1	Water	NA	NA	20.8	10.02	12/30/2021	18:02
4	ICV	1	Water	NA	NA	20.8	6.99	12/30/2021	18:05
5	CCV1	1	Water	NA	NA	20.6	2.01	12/30/2021	19:45
6	M5331-02	1	Solid	20.03	20	21.6	7.08	12/30/2021	19:46
7	M5340-01	1	Solid	20.04	20	21.4	11.64	12/30/2021	19:47
8	M5340-01DUP	1	Solid	20.03	20	21.5	11.65	12/30/2021	19:48
9	CCV2	1	Water	NA	NA	20.8	12.03	12/30/2021	19:50

WORKLIST(Hardcopy Internal Chain)

WB118031

WorkList Name : corrosivity m5336 s WorkList ID : 155652 Department : Wet-Chemistry Date : 12-30-2021 17:48:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/29/2021	Solid	M5331-02	Corrosivity	Cool 4 deg C	DVIR01	I11	WC-1	12/29/2021	9045D
12/29/2021	Solid	M5340-01	Corrosivity	Cool 4 deg C	ENTA05	I11	BAYSHORE-FILL	12/29/2021	9045D

Date/Time 12/30/21 16:50
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 12/30/21 20:00
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)