



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118032

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	12/30/2021	18:00
2	CAL2	1	Water	NA	NA	20.9	7.00	12/30/2021	18:01
3	CAL3	1	Water	NA	NA	20.8	10.02	12/30/2021	18:02
4	ICV	1	Water	NA	NA	20.8	6.99	12/30/2021	18:05
5	CCV1	1	Water	NA	NA	20.6	2.01	12/30/2021	18:06
6	M5327-03	1	Water	NA	NA	20.9	7.20	12/30/2021	18:09
7	M5327-04	1	Water	NA	NA	20.9	8.01	12/30/2021	18:10
8	M5327-05	1	Water	NA	NA	20.6	7.85	12/30/2021	18:13
9	M5327-06	1	Water	NA	NA	20.8	7.73	12/30/2021	18:15
10	M5327-07	1	Water	NA	NA	20.6	6.81	12/30/2021	18:16
11	M5327-08	1	Water	NA	NA	20.8	7.88	12/30/2021	18:20
12	M5327-09	1	Water	NA	NA	20.8	7.92	12/30/2021	18:23
13	M5328-01	1	Water	NA	NA	20.1	6.57	12/30/2021	18:25
14	M5328-02	1	Water	NA	NA	20.1	8.55	12/30/2021	18:26
15	M5328-03	1	Water	NA	NA	20.3	8.26	12/30/2021	18:30
16	CCV2	1	Water	NA	NA	20.9	12.03	12/30/2021	18:32
17	M5328-04	1	Water	NA	NA	20.6	6.19	12/30/2021	18:35
18	M5328-05	1	Water	NA	NA	20.6	6.24	12/30/2021	18:36
19	M5328-06	1	Water	NA	NA	20.6	7.70	12/30/2021	18:39
20	M5328-07	1	Water	NA	NA	20.9	6.15	12/30/2021	18:42

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M5328-08	1	Water	NA	NA	20.9	9.22	12/30/2021	18:45
22	M5328-09	1	Water	NA	NA	20.3	7.50	12/30/2021	18:46
23	M5328-10	1	Water	NA	NA	20.9	5.75	12/30/2021	18:50
24	M5328-11	1	Water	NA	NA	20.3	6.00	12/30/2021	18:51
25	M5335-01	1	Water	NA	NA	20.6	6.73	12/30/2021	18:52
26	M5335-02	1	Water	NA	NA	20.6	7.04	12/30/2021	18:53
27	CCV3	1	Water	NA	NA	20.7	2.01	12/30/2021	18:54
28	M5335-03	1	Water	NA	NA	20.4	6.56	12/30/2021	18:55
29	M5335-04	1	Water	NA	NA	20.1	6.94	12/30/2021	18:56
30	M5335-04DUP	1	Water	NA	NA	20.2	6.95	12/30/2021	18:57
31	CCV4	1	Water	NA	NA	20.1	12.03	12/30/2021	18:59

VB 118032

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m5335 w WorkList ID : 155653 Department : Wet-Chemistry Date : 12-30-2021 17:49:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/30/2021	Water	M5327-03	pH	Cool 4 deg C	PSEG03	I11	AUD-1235	12/30/2021	9040C
12/30/2021	Water	M5327-04	pH	Cool 4 deg C	PSEG03	I11	AUD-1236	12/30/2021	9040C
12/30/2021	Water	M5327-05	pH	Cool 4 deg C	PSEG03	I11	AUD-1237	12/30/2021	9040C
12/30/2021	Water	M5327-06	pH	Cool 4 deg C	PSEG03	I11	AUD-1238	12/30/2021	9040C
12/30/2021	Water	M5327-07	pH	Cool 4 deg C	PSEG03	I11	AUD-1239	12/30/2021	9040C
12/30/2021	Water	M5327-08	pH	Cool 4 deg C	PSEG03	I11	AUD-1240	12/30/2021	9040C
12/30/2021	Water	M5327-09	pH	Cool 4 deg C	PSEG03	I11	AUD-1241	12/30/2021	9040C
12/30/2021	Water	M5328-01	pH	Cool 4 deg C	PSEG03	I11	20938	12/30/2021	9040C
12/30/2021	Water	M5328-02	pH	Cool 4 deg C	PSEG03	I11	20936	12/30/2021	9040C
12/30/2021	Water	M5328-03	pH	Cool 4 deg C	PSEG03	I11	20948	12/30/2021	9040C
12/30/2021	Water	M5328-04	pH	Cool 4 deg C	PSEG03	I11	20994	12/30/2021	9040C
12/30/2021	Water	M5328-05	pH	Cool 4 deg C	PSEG03	I11	20992	12/30/2021	9040C
12/30/2021	Water	M5328-06	pH	Cool 4 deg C	PSEG03	I11	20993	12/30/2021	9040C
12/30/2021	Water	M5328-07	pH	Cool 4 deg C	PSEG03	I11	40440	12/30/2021	9040C
12/30/2021	Water	M5328-08	pH	Cool 4 deg C	PSEG03	I11	31320	12/30/2021	9040C
12/30/2021	Water	M5328-09	pH	Cool 4 deg C	PSEG03	I11	31321	12/30/2021	9040C
12/30/2021	Water	M5328-10	pH	Cool 4 deg C	PSEG03	I11	20995	12/30/2021	9040C
12/30/2021	Water	M5328-11	pH	Cool 4 deg C	PSEG03	I11	31322	12/30/2021	9040C
12/30/2021	Water	M5335-01	pH	Cool 4 deg C	PSEG03	I11	20968	12/30/2021	9040C
12/30/2021	Water	M5335-02	pH	Cool 4 deg C	PSEG03	I11	20967	12/30/2021	9040C
12/30/2021	Water	M5335-03	pH	Cool 4 deg C	PSEG03	I11	20966	12/30/2021	9040C
12/30/2021	Water	M5335-04	pH	Cool 4 deg C	PSEG03	I11	31276	12/30/2021	9040C

Date/Time: 12/30/21 17:50 JSP Date/Time: 12/30/21 JSP 18:30 19:00
 Raw Sample Received by: JSP (SM) Raw Sample Received by: JSP (SM)
 Raw Sample Relinquished by: JSP (WC) Raw Sample Relinquished by: JSP (WC)