



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118033

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	12/30/2021	18:00
2	CAL2	1	Water	NA	NA	20.9	7.00	12/30/2021	18:01
3	CAL3	1	Water	NA	NA	20.9	10.02	12/30/2021	18:02
4	ICV	1	Water	NA	NA	20.8	6.99	12/30/2021	18:05
5	CCV1	1	Water	NA	NA	20.6	2.01	12/30/2021	18:06
6	M5336-01	1	Solid	20.03	20	21.1	5.62	12/30/2021	19:05
7	M5336-03	1	Solid	20.04	20	21.2	6.17	12/30/2021	19:06
8	M5336-05	1	Solid	20.03	20	21.1	5.57	12/30/2021	19:10
9	M5336-07	1	Solid	20.04	20	21.2	7.83	12/30/2021	19:15
10	M5336-09	1	Solid	20.03	20	21.3	9.09	12/30/2021	19:16
11	M5336-11	1	Solid	20.04	20	21.2	9.19	12/30/2021	19:25
12	M5336-13	1	Solid	20.03	20	21.1	8.86	12/30/2021	19:35
13	M5336-15	1	Solid	20.04	20	21.4	9.04	12/30/2021	19:36
14	M5336-17	1	Solid	20.03	20	21.6	5.56	12/30/2021	18:39
15	M5336-17DUP	1	Solid	20.04	20	21.7	5.57	12/30/2021	19:40
16	CCV2	1	Water	NA	NA	20.8	12.02	12/30/2021	19:45

20118033

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m53367s

WorkList ID : 155654

Department : Wet-Chemistry

Date : 12-30-2021 17:50:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/30/2021	Solid	M5336-01	pH	Cool 4 deg C	PSEG03	I11	40569	12/30/2021	9045D
12/30/2021	Solid	M5336-03	pH	Cool 4 deg C	PSEG03	I11	40563	12/30/2021	9045D
12/30/2021	Solid	M5336-05	pH	Cool 4 deg C	PSEG03	I11	40566	12/30/2021	9045D
12/30/2021	Solid	M5336-07	pH	Cool 4 deg C	PSEG03	I11	40570	12/30/2021	9045D
12/30/2021	Solid	M5336-09	pH	Cool 4 deg C	PSEG03	I11	40564	12/30/2021	9045D
12/30/2021	Solid	M5336-11	pH	Cool 4 deg C	PSEG03	I11	40565	12/30/2021	9045D
12/30/2021	Solid	M5336-13	pH	Cool 4 deg C	PSEG03	I11	40567	12/30/2021	9045D
12/30/2021	Solid	M5336-15	pH	Cool 4 deg C	PSEG03	I11	40568	12/30/2021	9045D
12/30/2021	Solid	M5336-17	pH	Cool 4 deg C	PSEG03	I11	40562	12/30/2021	9045D

Date/Time 12/30/21 17:50
 Raw Sample Received by: JP (SM)
 Raw Sample Relinquished by: JP (SM)

Date/Time 12/30/21 18:30
 Raw Sample Received by: JP (SM)
 Raw Sample Relinquished by: JP (SM)