



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118048

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	01/03/2022	09:25
2	M5328-01		1	15	100.00	>212.0	>212.0	01/03/2022	09:40
3	M5328-01DUP		1	15	100.00	>212.0	>212.0	01/03/2022	10:05
4	M5328-02		1	12	100.00	>212.0	>212.0	01/03/2022	10:32
5	M5328-03		1	14	100.00	>212.0	>212.0	01/03/2022	10:58
6	M5328-04		1	13	100.00	>212.0	>212.0	01/03/2022	11:15
7	M5328-05		1	12	100.00	>212.0	>212.0	01/03/2022	11:40
8	M5328-06		1	13	100.00	>212.0	>212.0	01/03/2022	12:05
9	M5328-07		1	15	100.00	>212.0	>212.0	01/03/2022	12:30
10	M5328-08		1	14	100.00	>212.0	>212.0	01/03/2022	12:55
11	M5328-09		1	12	100.00	>212.0	>212.0	01/03/2022	13:20
12	M5328-10		1	15	100.00	>212.0	>212.0	01/03/2022	13:45
13	M5328-11		1	13	100.00	>212.0	>212.0	01/03/2022	14:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 118048

WorkList Name : fp123021 WorkList ID : 155637 Department : Wet-Chemistry Date : 12-30-2021 08:45:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/04/2022	Water	M5328-01	Flash Point	Cool 4 deg C	PSEG03	I11	20938	12/30/2021	1010B
01/04/2022	Water	M5328-02	Flash Point	Cool 4 deg C	PSEG03	I11	20936	12/30/2021	1010B
01/04/2022	Water	M5328-03	Flash Point	Cool 4 deg C	PSEG03	I11	20948	12/30/2021	1010B
01/04/2022	Water	M5328-04	Flash Point	Cool 4 deg C	PSEG03	I11	20994	12/30/2021	1010B
01/04/2022	Water	M5328-05	Flash Point	Cool 4 deg C	PSEG03	I11	20992	12/30/2021	1010B
01/04/2022	Water	M5328-06	Flash Point	Cool 4 deg C	PSEG03	I11	20993	12/30/2021	1010B
01/04/2022	Water	M5328-07	Flash Point	Cool 4 deg C	PSEG03	I11	40440	12/30/2021	1010B
01/04/2022	Water	M5328-08	Flash Point	Cool 4 deg C	PSEG03	I11	31320	12/30/2021	1010B
01/04/2022	Water	M5328-09	Flash Point	Cool 4 deg C	PSEG03	I11	31321	12/30/2021	1010B
01/04/2022	Water	M5328-10	Flash Point	Cool 4 deg C	PSEG03	I11	20995	12/30/2021	1010B
01/04/2022	Water	M5328-11	Flash Point	Cool 4 deg C	PSEG03	I11	31322	12/30/2021	1010B

Date/Time 01/03/22 09:05
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: RS (Ext-Lab)

Date/Time 01/03/22 14:10
Raw Sample Received by: RS (Ext-Lab)
Raw Sample Relinquished by: RS (WC)