

LB 118050


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44364



A= 0.0000e+000

B= 2.2684e-004

C= -8.6938e-002

Rho= 0.9997187

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.056	-0.056	135	0.000	135					
0.2	0.200	0.216	0.016	1337	0.0 %	1337					
2.5	2.500	2.628	0.128	11970	0.0 %	11970					
5.0	5.000	4.859	-0.141	21806	0.0 %	21806					
7.5	7.500	7.576	0.076	33781	0.0 %	33781					
10.0	10.000	9.976	-0.024	44364	0.0 %	44364					

0/00
8
5
-3
1
0

LB118050 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Jan 2022 08:43:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	135	-	0.0000	7471B	PPB	04 Jan 2022 08:54:15	S	Std
0.2 - 1	50.2	1337	-	0.2000	7471B	PPB	04 Jan 2022 08:59:01	S	Std
2.5 - 1	52.5	11970	-	2.5000	7471B	PPB	04 Jan 2022 09:01:18	S	Std
5.0 - 1	55	21806	-	5.0000	7471B	PPB	04 Jan 2022 09:06:06	S	Std
7.5 - 1	57.5	33781	-	7.5000	7471B	PPB	04 Jan 2022 09:13:30	S	Std
10.0 - 1	510	44364	-	10.0000	7471B	PPB	04 Jan 2022 09:23:32	S	Std
ICV31 - 1	ICV31	16678	3.6962	-	7471B	PPB	04 Jan 2022 09:30:08	U	SMPL
ICB31 - 1	ICB31	120	-0.0597	-	7471B	PPB	04 Jan 2022 09:32:23	U	SMPL
CCV37 - 1	CCV37	22069	4.9191	-	7471B	PPB	04 Jan 2022 09:34:38	U	SMPL
CCB37 - 1	CCB37	104	-0.0633	-	7471B	PPB	04 Jan 2022 09:36:53	U	SMPL
CRA - 1	CRA	1323	0.2132	-	7471B	PPB	04 Jan 2022 09:39:08	U	SMPL
HighStd - 1	HighStd	44839	10.0842	-	7471B	PPB	04 Jan 2022 09:41:23	U	SMPL
ChkStd - 1	ChkStd	31122	6.9727	-	7471B	PPB	04 Jan 2022 09:43:39	U	SMPL
PB141817BL - 1	PBS	266	-0.0266	-	7471B	PPB	04 Jan 2022 09:45:56	U	SMPL
PB141817BS - 1	LCSS	19496	4.3355	-	7471B	PPB	04 Jan 2022 09:48:12	U	SMPL
M5329-01 - 1	OR-3-123021	1775	0.3157	-	7471B	PPB	04 Jan 2022 09:55:36	U	SMPL
M5329-01DUP - 1	OR-3-123021DUP	1740	0.3078	-	7471B	PPB	04 Jan 2022 09:57:52	U	SMPL
M5329-01MS - 1	OR-3-123021MS	21293	4.7431	-	7471B	PPB	04 Jan 2022 10:00:09	U	SMPL
M5329-01MSD - 1	OR-3-123021MSD	21443	4.7771	-	7471B	PPB	04 Jan 2022 10:02:26	U	SMPL
M5337-09 - 1	WC-1	44489	10.0048	-	7471B	PPB	04 Jan 2022 10:04:42	U	SMPL
CCV38 - 1	CCV38	21974	4.8976	-	7471B	PPB	04 Jan 2022 10:06:57	U	SMPL
CCB38 - 1	CCB38	107	-0.0627	-	7471B	PPB	04 Jan 2022 10:09:14	U	SMPL
M5338-04 - 1	SB-4(8.5-9)	3884	0.7941	-	7471B	PPB	04 Jan 2022 10:11:30	U	SMPL
M5340-01 - 1	BAYSHORE-FILL	7281	1.5647	-	7471B	PPB	04 Jan 2022 10:13:45	U	SMPL
N1002-01 - 1	BLACK-DRUM	2170	0.4053	-	7471B	PPB	04 Jan 2022 10:16:01	U	SMPL
N1002-03 - 1	BLUE-DRUM	2026	0.3726	-	7471B	PPB	04 Jan 2022 10:18:16	U	SMPL
N1004-01 - 1	RO-R2-IEC	314031	HIGH	-	7471B	PPB	04 Jan 2022 10:20:32	X	
M5329-01LX5 - 1		233	-0.0341	-	7471B	PPB	04 Jan 2022 10:22:23	U	SMPL
M5329-01A - 1		20866	4.6462	-	7471B	PPB	04 Jan 2022 10:24:41	U	SMPL
N1004-01DLX10 - 1		43892	9.8694	-	7471B	PPB	04 Jan 2022 10:26:58	U	SMPL
M5337-09DLX10 - 1		22009	4.9055	-	7471B	PPB	04 Jan 2022 10:40:27	U	SMPL
CCV39 - 1	CCV39	22008	4.9053	-	7471B	PPB	04 Jan 2022 10:42:42	U	SMPL
CCB39 - 1	CCB39	252	-0.0298	-	7471B	PPB	04 Jan 2022 10:44:58	U	SMPL

MB
1/4/22