



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118055

Slope : 98.6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/04/2022	11:10
2	CAL2	1	Water	NA	NA	20.6	7.00	01/04/2022	11:11
3	CAL3	1	Water	NA	NA	20.6	10.03	01/04/2022	11:12
4	ICV	1	Water	NA	NA	20.7	7.00	01/04/2022	11:15
5	CCV1	1	Water	NA	NA	20.6	2.01	01/04/2022	11:16
6	N1011-01	1	Water	NA	NA	20.4	4.89	01/04/2022	11:20
7	N1011-01DUP	1	Water	NA	NA	20.5	4.90	01/04/2022	11:21
8	CCV2	1	Water	NA	NA	20.6	12.03	01/04/2022	11:25

VB 118055

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1011 WorkList ID : 155704 Department : Wet-Chemistry Date : 01-04-2022 10:57:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/30/2021	Water	N1011-01	pH	Cool 4 deg C	STER14	H11	JS10000	12/30/2021	9040C

Date/Time 01/04/22 11:00
Raw Sample Received by: JN (elf)
Raw Sample Relinquished by: JN (elf)

Date/Time 01/04/22 12:00
Raw Sample Received by: JN (elf)
Raw Sample Relinquished by: JN (elf)