

LB118056


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40797



A= 0.0000e+000

B= 2.4434e-004

C= 7.2516e-002

Rho= 0.9998171

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	0.079	0.079	28	0.000	28					-
0.2	0.200	0.227	0.027	631	0.0 %	631					14
2.5	2.500	2.358	-0.142	9354	0.0 %	9354					-6
5.0	5.000	5.021	0.021	20251	0.0 %	20251					0
7.5	7.500	7.475	-0.025	30295	0.0 %	30295					0
10.0	10.000	10.041	0.041	40797	0.0 %	40797					0

LB118056 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Jan 2022 12:51:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	28	-	0.0000	7470A	PPB	04 Jan 2022 13:31:04	S	Std
0.2 - 1	50.2	631	-	0.2000	7470A	PPB	04 Jan 2022 13:35:24	S	Std
2.5 - 1	52.5	9354	-	2.5000	7470A	PPB	04 Jan 2022 13:37:40	S	Std
5.0 - 1	55	20251	-	5.0000	7470A	PPB	04 Jan 2022 13:42:27	S	Std
7.5 - 1	57.5	30295	-	7.5000	7470A	PPB	04 Jan 2022 13:47:24	S	Std
10.0 - 1	510	40797	-	10.0000	7470A	PPB	04 Jan 2022 13:49:40	S	Std
ICV32 - 1	ICV32	15783	3.9289	-7470A	PPB	04 Jan 2022 13:52:22	U	SMPL	
ICB32 - 1	ICB32	-38	0.0632	-7470A	PPB	04 Jan 2022 13:54:37	U	SMPL	
CCV40 - 1	CCV40	21184	5.2485	-7470A	PPB	04 Jan 2022 13:56:53	U	SMPL	
CCB40 - 1	CCB40	-40	0.0627	-7470A	PPB	04 Jan 2022 13:59:09	U	SMPL	
CRA - 1	CRA	607	0.2208	-7470A	PPB	04 Jan 2022 14:01:24	U	SMPL	
HighStd - 1	HighStd	38038	9.3666	-7470A	PPB	04 Jan 2022 14:03:39	U	SMPL	
ChkStd - 1	ChkStd	28268	6.9794	-7470A	PPB	04 Jan 2022 14:05:55	U	SMPL	
PB141809BL - 1	PBW	-20	0.0676	-7470A	PPB	04 Jan 2022 14:08:12	U	SMPL	
PB141809BS - 1	LCSW	15834	3.9413	-7470A	PPB	04 Jan 2022 14:10:28	U	SMPL	
M5331-02 - 1	WC-1	-11	0.0698	-7470A	PPB	04 Jan 2022 14:12:44	U	SMPL	
M5336-01 - 1	40569	35	0.0811	-7470A	PPB	04 Jan 2022 14:15:01	U	SMPL	
M5336-03 - 1	40563	67	0.0889	-7470A	PPB	04 Jan 2022 14:17:18	U	SMPL	
M5336-05 - 1	40566	76	0.0911	-7470A	PPB	04 Jan 2022 14:19:35	U	SMPL	
M5336-07 - 1	40570	75	0.0908	-7470A	PPB	04 Jan 2022 14:21:51	U	SMPL	
CCV41 - 1	CCV41	21036	5.2124	-7470A	PPB	04 Jan 2022 14:24:07	U	SMPL	
CCB41 - 1	CCB41	-40	0.0627	-7470A	PPB	04 Jan 2022 14:26:22	U	SMPL	
M5336-09 - 1	40564	59	0.0869	-7470A	PPB	04 Jan 2022 14:28:37	U	SMPL	
M5336-11 - 1	40565	31	0.0801	-7470A	PPB	04 Jan 2022 14:30:52	U	SMPL	
M5336-13 - 1	40567	49	0.0845	-7470A	PPB	04 Jan 2022 14:33:08	U	SMPL	
M5336-15 - 1	40568	83	0.0928	-7470A	PPB	04 Jan 2022 14:35:23	U	SMPL	
M5336-17 - 1	40562	33	0.0806	-7470A	PPB	04 Jan 2022 14:37:40	U	SMPL	
M5336-17DUP - 1	40562DUP	44	0.0833	-7470A	PPB	04 Jan 2022 14:39:56	U	SMPL	
M5336-17MS - 1	40562MS	15196	3.7854	-7470A	PPB	04 Jan 2022 14:42:12	U	SMPL	
M5336-17MSD - 1	40562MSD	14536	3.6242	-7470A	PPB	04 Jan 2022 14:44:29	U	SMPL	
PB141810BL - 1	PBW	-9	0.0703	-7470A	PPB	04 Jan 2022 14:46:46	U	SMPL	
PB141810BS - 1	LCSW	13357	3.3361	-7470A	PPB	04 Jan 2022 14:49:01	U	SMPL	
CCV42 - 1	CCV42	20852	5.1674	-7470A	PPB	04 Jan 2022 14:51:17	U	SMPL	
CCB42 - 1	CCB42	-22	0.0671	-7470A	PPB	04 Jan 2022 14:53:33	U	SMPL	
M5327-01 - 1	AUD-1235-1241	20	0.0774	-7470A	PPB	04 Jan 2022 14:55:50	U	SMPL	
M5327-01DUP - 1	AUD-1235-1241DUP	21	0.0776	-7470A	PPB	04 Jan 2022 14:58:05	U	SMPL	
M5327-01MS - 1	AUD-1235-1241MS	15200	3.7864	-7470A	PPB	04 Jan 2022 15:00:21	U	SMPL	
M5327-01MSD - 1	AUD-1235-1241MSD	15336	3.8197	-7470A	PPB	04 Jan 2022 15:02:36	U	SMPL	
PB141809TB - 1	PB141782TB	13	0.0757	-7470A	PPB	04 Jan 2022 15:04:52	U	SMPL	
PB141810TB - 1	PB141787TB	18	0.0769	-7470A	PPB	04 Jan 2022 15:07:08	U	SMPL	
M5336-17LX5 - 1		66	0.0886	-7470A	PPB	04 Jan 2022 15:11:58	U	SMPL	
M5336-17A - 1		15990	3.9794	-7470A	PPB	04 Jan 2022 15:14:15	U	SMPL	
M5327-01LX5 - 1		-18	0.0681	-7470A	PPB	04 Jan 2022 15:16:32	U	SMPL	
M5327-01A - 1		16182	4.0264	-7470A	PPB	04 Jan 2022 15:18:48	U	SMPL	
CCV43 - 1	CCV43	21007	5.2053	-7470A	PPB	04 Jan 2022 15:21:03	U	SMPL	
CCB43 - 1	CCB43	-13	0.0693	-7470A	PPB	04 Jan 2022 15:23:19	U	SMPL	