



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118064

Slope : 98.5

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/04/2022	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/04/2022	16:11
3	CAL3	1	Water	NA	NA	20.3	10.03	01/04/2022	16:12
4	ICV	1	Water	NA	NA	20.4	6.98	01/04/2022	16:15
5	CCV1	1	Water	NA	NA	20.4	2.01	01/04/2022	16:58
6	N1007-03	1	Solid	20.03	20	20.6	11.44	01/04/2022	17:00
7	N1014-02	1	Solid	20.04	20	20.6	7.97	01/04/2022	17:09
8	N1016-02	1	Solid	20.02	20	20.6	9.34	01/04/2022	17:15
9	N1016-02DUP	1	Solid	20.03	20	20.7	9.35	01/04/2022	17:16
10	CCV2	1	Water	NA	NA	20.5	12.03	01/04/2022	17:20

LB 118064

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1007 WorkList ID : 155715 Department : Wet-Chemistry Date : 01-04-2022 13:49:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/04/2022	Solid	N1007-03	Corrosivity	Cool 4 deg C	TULL02	I11	BARNUM-ST-SP	01/04/2022	9045D
01/04/2022	Solid	N1014-02	Corrosivity	Cool 4 deg C	PSEG03	I11	72-11944	01/04/2022	9045D
01/04/2022	Solid	N1016-02	Corrosivity	Cool 4 deg C	ENTA05	I11	E3	01/04/2022	9045D

Date/Time 01/04/22
Raw Sample Received by: JH (WCC)
Raw Sample Relinquished by: JT SM

Date/Time 01/04/22
Raw Sample Received by: JT SM
Raw Sample Relinquished by: JH (WCC)