



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118070

Slope : 98.4

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/05/2022	11:00
2	CAL2	1	Water	NA	NA	20.6	7.00	01/05/2022	11:01
3	CAL3	1	Water	NA	NA	20.6	10.03	01/05/2022	11:02
4	ICV	1	Water	NA	NA	20.7	7.01	01/05/2022	11:05
5	CCV1	1	Water	NA	NA	20.6	2.01	01/05/2022	11:06
6	N1018-01	1	Solid	20.03	20	20.1	8.23	01/05/2022	11:10
7	N1018-04	1	Solid	20.04	20	20.3	8.11	01/05/2022	11:15
8	N1018-06	1	Solid	20.02	20	20.4	7.49	01/05/2022	11:16
9	N1018-06DUP	1	Solid	20.03	20	20.5	7.50	01/05/2022	11:17
10	CCV2	1	Water	NA	NA	20.7	12.03	01/05/2022	11:20

W 118010

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1018 WorkList ID : 155741 Department : Wet-Chemistry Date : 01-05-2022 10:40:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/04/2022	Solid	N1018-01	pH	Cool 4 deg C	WEST10	D11	FM-A4-41-01-4.5	01/04/2022	9045D
01/04/2022	Solid	N1018-04	pH	Cool 4 deg C	WEST10	D11	DUP-01-010422	01/04/2022	9045D
01/04/2022	Solid	N1018-06	pH	Cool 4 deg C	WEST10	D11	FM-E3-02-4.5	01/04/2022	9045D

22.05.2022
11.05.2022
01.05.2022

Date/Time 01/05/22 10:45
Raw Sample Received by: JN (WCC)
Raw Sample Relinquished by: RS (EAT-106)

Date/Time 01/05/22 12:10
Raw Sample Received by: RS (EAT-106)
Raw Sample Relinquished by: JN (WCC)