



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118080

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	01/05/2022	11:25
2	N1012-02		1	12	100.00	>212.0	>212.0	01/05/2022	11:40
3	N1012-02DUP		1	12	100.00	>212.0	>212.0	01/05/2022	12:05
4	N1012-14		1	15	100.00	>212.0	>212.0	01/05/2022	12:35
5	N1012-18		1	14	100.00	>212.0	>212.0	01/05/2022	12:58
6	N1012-19		1	13	100.00	>212.0	>212.0	01/05/2022	13:27
7	N1012-20		1	15	100.00	>212.0	>212.0	01/05/2022	13:45
8	N1012-21		1	14	100.00	>212.0	>212.0	01/05/2022	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 112080

WorkList Name : FP1522 WorkList ID : 155736 Department : Wet-Chemistry Date : 01-05-2022 09:44:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/11/2022	Water	N1012-02	Flash Point	Cool 4 deg C	PSEG03	I11	31212	01/04/2022	1010B
01/11/2022	Water	N1012-14	Flash Point	Cool 4 deg C	PSEG03	I11	20989	01/04/2022	1010B
01/11/2022	Water	N1012-18	Flash Point	Cool 4 deg C	PSEG03	I11	COMP-31235:38	01/04/2022	1010B
01/11/2022	Water	N1012-19	Flash Point	Cool 4 deg C	PSEG03	I11	COMP-31239:42	01/04/2022	1010B
01/11/2022	Water	N1012-20	Flash Point	Cool 4 deg C	PSEG03	I11	COMP-31243:46	01/04/2022	1010B
01/11/2022	Water	N1012-21	Flash Point	Cool 4 deg C	PSEG03	I11	COMP-31332:34	01/04/2022	1010B

Date/Time 01/05/22 09:50
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 01/05/22 14:15
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)