



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118084

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/07/2022	08:25
2	CAL2	1	Water	NA	NA	20.6	7.00	01/06/2022	08:26
3	CAL3	1	Water	NA	NA	20.6	10.03	01/06/2022	08:27
4	ICV	1	Water	NA	NA	20.6	7.02	01/06/2022	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	01/06/2022	08:32
6	N1023-01	1	Solid	20.03	20	20.9	7.17	01/06/2022	08:49
7	N1023-01DUP	1	Solid	20.04	20	20.8	7.18	01/06/2022	08:50
8	CCV2	1	Water	NA	NA	20.3	12.03	01/06/2022	08:51

27118084

WORKLIST(Hardcopy Internal Chain)

CORROSIVITY N1036 WorkList ID : 155755 Department : Wet-Chemistry Date : 01-06-2022

Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	M
Solid	N1023-01	Corrosivity	Cool 4 deg C	PSEG03	I11	HH-DRUM-COMP	01/05/2022	9