



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118085

Slope : 98.8

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/06/2022	09:00
2	CAL2	1	Water	NA	NA	20.3	7.00	01/06/2022	09:01
3	CAL3	1	Water	NA	NA	20.4	10.03	01/06/2022	09:02
4	ICV	1	Water	NA	NA	20.5	7.00	01/06/2022	09:05
5	CCV1	1	Water	NA	NA	20.5	2.01	01/06/2022	09:06
6	N1034-01	1	Water	NA	NA	20.1	7.95	01/06/2022	09:11
7	N1034-01DUP	1	Water	NA	NA	20.2	7.96	01/06/2022	09:12
8	N1037-03	1	Water	NA	NA	20.2	7.65	01/06/2022	09:15
9	N1037-05	1	Water	NA	NA	20.4	7.38	01/06/2022	09:20
10	N1047-05	1	Water	NA	NA	20.1	7.70	01/06/2022	09:22
11	CCV2	1	Water	NA	NA	20.5	12.03	01/06/2022	09:25