


LB 118086

7471B

INSTRUMENT ID: CV2

 μ Abs.:

53404



Linear

A= 0.0000e+000

B= 1.8804e-004

C= -1.4700e-001

Rho= 0.9998252

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.056	-0.056	483	0.000	483					7
0.2	0.200	0.215	0.015	1927	0.0 %	1927					-1
2.5	2.500	2.482	-0.018	13981	0.0 %	13981					-1
5.0	5.000	5.066	0.066	27724	0.0 %	27724					-1
7.5	7.500	7.597	0.097	41184	0.0 %	41184					-1
10.0	10.000	9.895	-0.105	53404	0.0 %	53404					-1

LB118086 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jan 2022 10:01:14

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	483	-	0.0000	7471B	PPB	06 Jan 2022 10:45:12	S	Std
0.2 - 1	50.2	1927	-	0.2000	7471B	PPB	06 Jan 2022 10:47:30	S	Std
2.5 - 1	52.5	13981	-	2.5000	7471B	PPB	06 Jan 2022 10:49:47	S	Std
5.0 - 1	55	27724	-	5.0000	7471B	PPB	06 Jan 2022 10:52:05	S	Std
7.5 - 1	57.5	41184	-	7.5000	7471B	PPB	06 Jan 2022 10:54:22	S	Std
10.0 - 1	60	53404	-	10.0000	7471B	PPB	06 Jan 2022 11:04:22	S	Std
ICV34 - 1	ICV34	22143	4.0168	-	7471B	PPB	06 Jan 2022 11:07:06	U	SMPL
ICB34 - 1	ICB34	1016	0.0441	-	7471B	PPB	06 Jan 2022 11:09:21	U	SMPL
CCV49 - 1	CCV49	27610	5.0448	-	7471B	PPB	06 Jan 2022 11:11:37	U	SMPL
CCB49 - 1	CCB49	1012	0.0433	-	7471B	PPB	06 Jan 2022 11:13:52	U	SMPL
CRA - 1	CRA	1882	0.2069	-	7471B	PPB	06 Jan 2022 11:16:08	U	SMPL
HighStd - 1	HighStd	53848	9.9787	-	7471B	PPB	06 Jan 2022 11:18:23	U	SMPL
ChkStd - 1	ChkStd	37851	6.9706	-	7471B	PPB	06 Jan 2022 11:20:38	U	SMPL
PB141858BL - 1	PBS	527	-0.0479	-	7471B	PPB	06 Jan 2022 11:25:30	U	SMPL
PB141858BS - 1	LCSS	20450	3.6985	-	7471B	PPB	06 Jan 2022 11:27:46	U	SMPL
N1007-02 - 1	BARNUM-ST-SP	7407	1.2458	-	7471B	PPB	06 Jan 2022 11:30:02	U	SMPL
N1014-01 - 1	72-11944	2964	0.4104	-	7471B	PPB	06 Jan 2022 11:32:19	U	SMPL
N1016-01 - 1	E3	458951	86.1551	-	7471B	PPB	06 Jan 2022 11:34:36	U	SMPL
N1018-01 - 1	FM-A4-41-01-4.5	1648	0.1629	-	7471B	PPB	06 Jan 2022 11:36:53	U	SMPL
N1018-04 - 1	DUP-01-010422	4819	0.7592	-	7471B	PPB	06 Jan 2022 11:39:29	U	SMPL
CCV50 - 1	CCV50	26071	4.7554	-	7471B	PPB	06 Jan 2022 11:41:44	U	SMPL
CCB50 - 1	CCB50	923	0.0266	-	7471B	PPB	06 Jan 2022 11:44:00	U	SMPL
N1018-06 - 1	FM-E3-02-4.5	2580	0.3381	-	7471B	PPB	06 Jan 2022 11:51:02	U	SMPL
N1018-06DUP - 1	FM-E3-02-4.5DUP	2379	0.3004	-	7471B	PPB	06 Jan 2022 11:53:17	U	SMPL
N1018-07 - 1	N1018-06MS	21529	3.9014	-	7471B	PPB	06 Jan 2022 11:55:32	U	SMPL
N1018-08 - 1	N1018-06MSD	20565	3.7201	-	7471B	PPB	06 Jan 2022 11:57:47	U	SMPL
N1029-02 - 1	ETGI-366	1434	0.1227	-	7471B	PPB	06 Jan 2022 12:00:03	U	SMPL
N1038-01 - 1	RP-RB21137	2838	0.3867	-	7471B	PPB	06 Jan 2022 12:02:19	U	SMPL
N1039-01 - 1	S-1	2465	0.3165	-	7471B	PPB	06 Jan 2022 12:04:36	U	SMPL
N1047-01 - 1	FM-E8-01-4.5	5411	0.8705	-	7471B	PPB	06 Jan 2022 12:06:53	U	SMPL
N1047-03 - 1	FM-E7-01-2.5	8149	1.3854	-	7471B	PPB	06 Jan 2022 12:09:10	U	SMPL
N1047-04 - 1	FM-E2-01-4.0	5253	0.8408	-	7471B	PPB	06 Jan 2022 12:11:26	U	SMPL
CCV51 - 1	CCV51	26972	4.9249	-	7471B	PPB	06 Jan 2022 12:13:41	U	SMPL
CCB51 - 1	CCB51	992	0.0395	-	7471B	PPB	06 Jan 2022 12:15:57	U	SMPL
N1018-06LX5 - 1		854	0.0136	-	7471B	PPB	06 Jan 2022 12:18:13	U	SMPL
N1018-06A - 1		21278	3.8542	-	7471B	PPB	06 Jan 2022 12:20:29	U	SMPL
N1016-01DLX25 - 1		23420	4.2569	-	7471B	PPB	06 Jan 2022 12:28:48	U	SMPL
CCV52 - 1	CCV52	26698	4.8733	-	7471B	PPB	06 Jan 2022 12:31:06	U	SMPL
CCB52 - 1	CCB52	966	0.0346	-	7471B	PPB	06 Jan 2022 12:33:21	U	SMPL