



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB118092

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	01/06/2022	09:32
2	N1015-02		1	15	100.00	>212.0	>212.0	01/06/2022	09:58
3	N1015-02DUP		1	15	100.00	>212.0	>212.0	01/06/2022	10:20
4	N1015-03		1	12	100.00	>212.0	>212.0	01/06/2022	10:45
5	N1015-04		1	13	100.00	>212.0	>212.0	01/06/2022	11:10
6	N1015-05		1	14	100.00	>212.0	>212.0	01/06/2022	11:35
7	N1015-06		1	11	100.00	>212.0	>212.0	01/06/2022	11:55
8	N1015-07		1	12	100.00	>212.0	>212.0	01/06/2022	12:15
9	N1015-08		1	15	100.00	>212.0	>212.0	01/06/2022	12:35
10	N1015-09		1	14	100.00	>212.0	>212.0	01/06/2022	12:58
11	N1015-10		1	13	100.00	>212.0	>212.0	01/06/2022	13:22
12	N1037-03		1	15	100.00	>212.0	>212.0	01/06/2022	13:45
13	N1037-05		1	14	100.00	>212.0	>212.0	01/06/2022	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP1522-1 WorkList ID : 155737 Department : Wet-Chemistry Date : 01-05-2022 09:46:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/11/2022	Water	N1015-02	Flash Point	Cool 4 deg C	PSEG03	I11	30915	01/04/2022	1010B
01/11/2022	Water	N1015-03	Flash Point	Cool 4 deg C	PSEG03	I11	20905	01/04/2022	1010B
01/11/2022	Water	N1015-04	Flash Point	Cool 4 deg C	PSEG03	I11	20969	01/04/2022	1010B
01/11/2022	Water	N1015-05	Flash Point	Cool 4 deg C	PSEG03	I11	31282	01/04/2022	1010B
01/11/2022	Water	N1015-06	Flash Point	Cool 4 deg C	PSEG03	I11	20985	01/04/2022	1010B
01/11/2022	Water	N1015-07	Flash Point	Cool 4 deg C	PSEG03	I11	20986	01/04/2022	1010B
01/11/2022	Water	N1015-08	Flash Point	Cool 4 deg C	PSEG03	I11	20987	01/04/2022	1010B
01/11/2022	Water	N1015-09	Flash Point	Cool 4 deg C	PSEG03	I11	31278	01/04/2022	1010B
01/11/2022	Water	N1015-10	Flash Point	Cool 4 deg C	PSEG03	I11	31279	01/04/2022	1010B
01/10/2022	Water	N1037-03	Flash Point	Cool 4 deg C	PSEG03	I11	NWB-1542	01/05/2022	1010B
01/10/2022	Water	N1037-05	Flash Point	Cool 4 deg C	PSEG03	I11	NWB-1729	01/05/2022	1010B

Date/Time 01/06/22 09:05
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: RJCFE (146)

Date/Time 01/06/22 14:15
Raw Sample Received by: RJCFE (146)
Raw Sample Relinquished by: KS (WC)