



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118098

Slope : 99.1

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/06/2022	15:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/06/2022	15:11
3	CAL3	1	Water	NA	NA	20.3	10.03	01/06/2022	15:12
4	ICV	1	Water	NA	NA	20.4	7.02	01/06/2022	15:15
5	CCV1	1	Water	NA	NA	20.4	2.01	01/06/2022	15:16
6	N1057-01	1	Water	NA	NA	20.4	6.86	01/06/2022	15:19
7	N1057-01DUP	1	Water	NA	NA	20.5	6.87	01/06/2022	15:20
8	N1058-01	1	Water	NA	NA	20.3	6.92	01/06/2022	15:22
9	N1059-01	1	Water	NA	NA	20.5	4.81	01/06/2022	15:25
10	N1059-02	1	Water	NA	NA	20.4	4.88	01/06/2022	15:29
11	CCV2	1	Water	NA	NA	20.6	12.02	01/06/2022	15:30

VB 118098

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1067

WorkList ID : 155796

Department : Wet-Chemistry

Date : 01-06-2022 14:56:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2022	Water	N1057-01	A pH	Cool 4 deg C	PSEG03	I11	RSW003	01/06/2022	9040C
01/06/2022	Water	N1058-01	A pH	Cool 4 deg C	PSEG03	I21	DSN001	01/06/2022	9040C
12/31/2021	Water	N1059-01	B pH	Cool 4 deg C	STER16	H11	MAIN-CELL	12/31/2021	9040C
12/31/2021	Water	N1059-02	B pH	Cool 4 deg C	STER16	H11	EXCELL	12/31/2021	9040C

Date/Time 01/06/22 15:00
 Raw Sample Received by: JH (CSC)
 Raw Sample Relinquished by: RJ (CSC-1ab)

Date/Time 01/06/22 14:56
 Raw Sample Received by: RJ (CSC-1ab)
 Raw Sample Relinquished by: JH (CSC)