



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118103

Slope : 98.3

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/07/2022	08:20
2	CAL2	1	Water	NA	NA	20.3	7.00	01/07/2022	08:21
3	CAL3	1	Water	NA	NA	20.3	10.03	01/07/2022	08:22
4	ICV	1	Water	NA	NA	20.4	7.02	01/07/2022	08:25
5	CCV1	1	Water	NA	NA	20.4	2.01	01/07/2022	08:26
6	N1071-01	1	Solid	20.03	20	20.2	6.71	01/07/2022	08:30
7	N1071-01DUP	1	Solid	20.04	20	20.3	6.72	01/07/2022	08:31
8	N1071-02	1	Solid	20.02	20	20.3	6.53	01/07/2022	08:33
9	N1071-03	1	Solid	20.05	20	20.4	7.84	01/07/2022	08:35
10	CCV2	1	Water	NA	NA	20.4	12.03	01/07/2022	08:36

VB 118103

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S M1071 WorkList ID : 155804 Department : Wet-Chemistry Date : 01-07-2022 08:13:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2022	Solid	N1071-01	pH	Cool 4 deg C	WEST10	I21	FM-E6-01-4.5	01/06/2022	9045D
01/05/2022	Solid	N1071-02	pH	Cool 4 deg C	WEST10	I21	FM-E1-01-4.5	01/05/2022	9045D
01/06/2022	Solid	N1071-03	pH	Cool 4 deg C	WEST10	I21	FW-0226-01-4.5	01/06/2022	9045D

Date/Time 21/07/22 08:15
 Raw Sample Received by: JP Cooper
 Raw Sample Relinquished by: JP Cooper

Date/Time 21/07/22 11:00
 Raw Sample Received by: JP Cooper
 Raw Sample Relinquished by: JP Cooper