



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118104

Slope : 98.3

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/07/2022	08:20
2	CAL2	1	Water	NA	NA	20.3	7.00	01/07/2022	08:21
3	CAL3	1	Water	NA	NA	20.3	10.03	01/07/2022	08:22
4	ICV	1	Water	NA	NA	20.4	7.02	01/07/2022	08:25
5	CCV1	1	Water	NA	NA	20.4	2.01	01/07/2022	08:26
6	N1050-03	1	Water	NA	NA	20.6	7.21	01/07/2022	08:40
7	N1050-03DUP	1	Water	NA	NA	20.7	7.22	01/07/2022	08:41
8	N1071-04	1	Water	NA	NA	20.1	8.09	01/07/2022	08:43
9	CCV2	1	Water	NA	NA	20.4	12.02	01/07/2022	08:45

LB 118104

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH WM1071 WorkList ID : 155805 Department : Wet-Chemistry Date : 01-07-2022 08:13:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2022	Water	N1050-03	pH	Cool 4 deg C	PSEG03	I21	P1660	01/06/2022	9040C
01/06/2022	Water	N1071-04	pH	Cool 4 deg C	WEST10	I21	FB-010622	01/06/2022	9040C

Date/Time 01/07/22 08:15
Raw Sample Received by: JP (WC)
Raw Sample Relinquished by: CP SM

Date/Time 01/07/22 11:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP (WC)