



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118105

Slope : 98.9

BalanceID: WC SC-6

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/07/2022	09:25
2	CAL2	1	Water	NA	NA	20.6	7.00	01/07/2022	09:26
3	CAL3	1	Water	NA	NA	20.6	10.02	01/07/2022	09:27
4	ICV	1	Water	NA	NA	20.5	6.99	01/07/2022	09:30
5	CCV1	1	Water	NA	NA	20.6	2.01	01/07/2022	09:32
6	N1049-01	1	Solid	20.03	20	20.8	7.22	01/07/2022	09:35
7	N1049-01DUP	1	Solid	20.04	20	20.9	7.23	01/07/2022	09:36
8	N1049-03	1	Solid	20.02	20	20.4	8.78	01/07/2022	09:37
9	N1049-05	1	Solid	20.03	20	20.9	8.91	01/07/2022	09:39
10	N1052-02	1	Solid	20.02	20	20.3	11.20	01/07/2022	09:40
11	N1056-01	1	Solid	20.01	20	20.2	8.50	01/07/2022	09:42
12	N1056-02	1	Solid	20.02	20	20.4	8.05	01/07/2022	09:43
13	N1056-03	1	Solid	20.01	20	20.1	6.82	01/07/2022	09:45
14	N1056-04	1	Solid	20.03	20	20.5	7.21	01/07/2022	09:46
15	N1056-05	1	Solid	20.02	20	20.7	8.02	01/07/2022	09:50
16	CCV2	1	Water	NA	NA	20.4	12.03	01/07/2022	09:52
17	N1056-06	1	Solid	20.01	20	20.3	5.21	01/07/2022	10:00
18	N1065-01	1	Solid	20.03	20	20.6	5.74	01/07/2022	10:02
19	N1065-06	1	Solid	20.04	20	20.4	10.36	01/07/2022	10:06
20	N1065-10	1	Solid	20.03	20	20.3	8.01	01/07/2022	10:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N1065-14	1	Solid	20.04	20	20.1	7.53	01/07/2022	10:11
22	N1065-18	1	Solid	20.03	20	20.4	8.44	01/07/2022	10:15
23	N1065-22	1	Solid	20.04	20	20.4	6.75	01/07/2022	10:19
24	N1065-26	1	Solid	20.03	20	20.1	6.99	01/07/2022	10:22
25	N1065-30	1	Solid	20.04	20	20.7	7.32	01/07/2022	10:25
26	N1065-34	1	Solid	20.03	20	20.4	5.35	01/07/2022	10:26
27	CCV3	1	Water	NA	NA	20.3	2.01	01/07/2022	10:30
28	N1065-38	1	Solid	20.04	20	20.4	6.22	01/07/2022	10:32
29	N1065-42	1	Solid	20.02	20	20.4	5.87	01/07/2022	10:35
30	N1065-42DUP	1	Solid	20.03	20	20.5	5.88	01/07/2022	10:36
31	CCV4	1	Water	NA	NA	20.3	12.03	01/07/2022	10:39