



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB118106

Slope : 98.9

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	01/07/2022	09:25
2	CAL2	1	Water	NA	NA	20.6	7.00	01/07/2022	09:26
3	CAL3	1	Water	NA	NA	20.6	10.02	01/07/2022	09:27
4	ICV	1	Water	NA	NA	20.5	6.99	01/07/2022	09:30
5	CCV1	1	Water	NA	NA	20.6	2.01	01/07/2022	09:32
6	N1056-07	1	Water	NA	NA	20.8	7.75	01/07/2022	10:40
7	N1056-07DUP	1	Water	NA	NA	20.9	7.76	01/07/2022	10:41
8	N1056-08	1	Water	NA	NA	20.4	7.20	01/07/2022	10:45
9	CCV2	1	Water	NA	NA	20.4	12.03	01/07/2022	10:46

VP 118106

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N1056 W WorkList ID : 155831 Department : Wet-Chemistry Date : 01-07-2022 10:05:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2022	Water	N1056-07	Corrosivity	Cool 4 deg C	PSEG03	I21	COMP-3	01/06/2022	9040C
01/06/2022	Water	N1056-08	Corrosivity	Cool 4 deg C	PSEG03	I21	1917-WATER	01/06/2022	9040C

Date/Time 01/06/22 10:10
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 01/06/22 12:00
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JD (WC)