

LB118108


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39336



A= 0.0000e+000

B= 2.5691e-004

C= 2.8920e-003

Rho= 0.9997993

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/D
0.0	0.000	0.056	0.056	205	0.000	205					7
0.2	0.200	0.199	-0.001	762	0.0 %	762					-1
2.5	2.500	2.491	-0.009	9684	0.0 %	9684					0
5.0	5.000	4.978	-0.022	19363	0.0 %	19363					0
7.5	7.500	7.369	-0.131	28670	0.0 %	28670					2
10.0	10.000	10.109	0.109	39336	0.0 %	39336					1

LB118108 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jan 2022 10:55:11

Sample ID	Extended ID	μ Abs	Conc	Strnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	205	-	0.0000	7470A	PPB	07 Jan 2022 11:09:41	S	Std
0.2 - 1	51.2	762	-	0.2000	7470A	PPB	07 Jan 2022 11:18:16	S	Std
2.5 - 1	52.5	9684	-	2.5000	7470A	PPB	07 Jan 2022 11:20:34	S	Std
5.0 - 1	55	19363	-	5.0000	7470A	PPB	07 Jan 2022 11:22:51	S	Std
7.5 - 1	53.5	28670	-	7.5000	7470A	PPB	07 Jan 2022 11:25:07	S	Std
10.0 - 1	51.0	39336	-	10.0000	7470A	PPB	07 Jan 2022 11:37:02	S	Std
ICV35 - 1	ICV35	14141	3.6359	-	7470A	PPB	07 Jan 2022 11:44:49	U	SMPL
ICB35 - 1	ICB35	22	0.0085	-	7470A	PPB	07 Jan 2022 11:49:34	U	SMPL
CCV53 - 1	CCV53	19635	5.0474	-	7470A	PPB	07 Jan 2022 11:54:23	U	SMPL
CCB53 - 1	CCB53	-38	-0.0069	-	7470A	PPB	07 Jan 2022 11:56:38	U	SMPL
CRA - 1	CRA	754	0.1966	-	7470A	PPB	07 Jan 2022 11:58:53	U	SMPL
HighStd - 1	HighStd	37992	9.7635	-	7470A	PPB	07 Jan 2022 12:01:08	U	SMPL
ChkStd - 1	ChkStd	25454	6.5424	-	7470A	PPB	07 Jan 2022 12:11:33	U	SMPL
PB141885BL - 1	PBW	-56	-0.0115	-	7470A	PPB	07 Jan 2022 12:13:49	U	SMPL
PB141885BS - 1	LCSW	14834	3.8139	-	7470A	PPB	07 Jan 2022 12:16:05	U	SMPL
N1037-03 - 1	NWB-1542	0	0.0029	-	7470A	PPB	07 Jan 2022 12:18:21	U	SMPL
N1037-03DUP - 1	NWB-1542DUP	45	0.0145	-	7470A	PPB	07 Jan 2022 12:20:38	U	SMPL
N1037-03MS - 1	NWB-1542MS	13758	3.5375	-	7470A	PPB	07 Jan 2022 12:22:55	U	SMPL
N1037-03MSD - 1	NWB-1542MSD	13751	3.5357	-	7470A	PPB	07 Jan 2022 12:25:12	U	SMPL
N1037-05 - 1	NWB-1729	50	0.0157	-	7470A	PPB	07 Jan 2022 12:27:28	U	SMPL
CCV54 - 1	CCV54	19397	4.9862	-	7470A	PPB	07 Jan 2022 12:29:43	U	SMPL
CCB54 - 1	CCB54	-26	-0.0038	-	7470A	PPB	07 Jan 2022 12:31:58	U	SMPL
PB141886BL - 1	PBW	51	0.0160	-	7470A	PPB	07 Jan 2022 12:34:13	U	SMPL
PB141886BS - 1	LCSW	13644	3.5082	-	7470A	PPB	07 Jan 2022 12:36:28	U	SMPL
N1037-01 - 1	NWB-1681	22	0.0085	-	7470A	PPB	07 Jan 2022 12:38:44	U	SMPL
N1037-01DUP - 1	NWB-1681DUP	55	0.0170	-	7470A	PPB	07 Jan 2022 12:41:00	U	SMPL
N1037-01MS - 1	NWB-1681MS	12558	3.2292	-	7470A	PPB	07 Jan 2022 12:43:16	U	SMPL
N1037-01MSD - 1	NWB-1681MSD	14767	3.7967	-	7470A	PPB	07 Jan 2022 12:45:33	U	SMPL
N1042-01 - 1	TRE-1268	-10	-0.0003	-	7470A	PPB	07 Jan 2022 12:47:50	U	SMPL
N1042-03 - 1	TRE-1522	-8	-0.0029	-	7470A	PPB	07 Jan 2022 12:50:07	U	SMPL
N1042-01 - 1	TRE-1268	319	0.0848	-	7470A	PPB	07 Jan 2022 12:51:18	U	SMPL
N1042-03 - 1	TRE-1522	-10	0.0003	-	7470A	PPB	07 Jan 2022 12:53:35	U	SMPL
N1044-01 - 1	TREMONT-AVE-STATION-DRUM-14-POLY	43	0.0139	-	7470A	PPB	07 Jan 2022 12:55:52	U	SMPL
N1044-02 - 1	TREMONT-AVE-STATION-DRUM-15-PPE	54	0.0168	-	7470A	PPB	07 Jan 2022 12:58:08	U	SMPL
CCV55 - 1	CCV55	19908	5.1175	-	7470A	PPB	07 Jan 2022 13:00:23	U	SMPL
CCB55 - 1	CCB55	-44	-0.0084	-	7470A	PPB	07 Jan 2022 13:02:39	U	SMPL
PB141889BL - 1	PBW	59	0.0180	-	7470A	PPB	07 Jan 2022 13:04:54	U	SMPL
PB141889BS - 1	LCSW	13297	3.4191	-	7470A	PPB	07 Jan 2022 13:07:10	U	SMPL
N1022-02 - 1	COMP	188	0.0512	-	7470A	PPB	07 Jan 2022 13:09:25	U	SMPL
N1041-01 - 1	DEWATERING-EFFLUENT	62	0.0188	-	7470A	PPB	07 Jan 2022 13:11:41	U	SMPL
N1041-01DUP - 1	DEWATERING-EFFLUENTDUP	82	0.0240	-	7470A	PPB	07 Jan 2022 13:13:56	U	SMPL
N1041-01MS - 1	DEWATERING-EFFLUENTMS	13326	3.4265	-	7470A	PPB	07 Jan 2022 13:16:13	U	SMPL
N1041-01MSD - 1	DEWATERING-EFFLUENTMSD	13228	3.4013	-	7470A	PPB	07 Jan 2022 13:18:30	U	SMPL
N1047-05 - 1	FB-010522	46	0.0147	-	7470A	PPB	07 Jan 2022 13:20:47	U	SMPL
PB141859TB - 1	PB141859TB	49	0.0155	-	7470A	PPB	07 Jan 2022 13:23:04	U	SMPL
PB141870TB - 1	PB141870TB	62	0.0188	-	7470A	PPB	07 Jan 2022 13:25:20	U	SMPL
CCV56 - 1	CCV56	20386	5.2403	-	7470A	PPB	07 Jan 2022 13:39:38	U	SMPL
CCB56 - 1	CCB56	-14	-0.0007	-	7470A	PPB	07 Jan 2022 13:44:08	U	SMPL
N1037-03LX5 - 1		-13	-0.0004	-	7470A	PPB	07 Jan 2022 13:46:23	U	SMPL
N1037-03A - 1		13032	3.3510	-	7470A	PPB	07 Jan 2022 13:48:38	U	SMPL
N1041-01LX5 - 1		3	0.0037	-	7470A	PPB	07 Jan 2022 13:50:54	U	SMPL
N1041-01A - 1		14599	3.7536	-	7470A	PPB	07 Jan 2022 13:53:09	U	SMPL
N1041-01LX5 - 1		-102	-0.0233	-	7470A	PPB	07 Jan 2022 13:55:25	U	SMPL
N1041-01A - 1		14892	3.8288	-	7470A	PPB	07 Jan 2022 13:57:42	U	SMPL
CCV57 - 1	CCV57	19309	4.9636	-	7470A	PPB	07 Jan 2022 14:13:40	U	SMPL
CCW57 - 1	CCW57	-147	-0.0349	-	7470A	PPB	07 Jan 2022 14:15:57	U	SMPL

CCB57

023

1/2/21