



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB118115

Slope : 98.9

pH Meter ID : WCPH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/07/2022	13:30
2	CAL2	1	Water	NA	NA	20.8	7.00	01/07/2022	13:31
3	CAL3	1	Water	NA	NA	20.8	10.03	01/07/2022	13:32
4	ICV	1	Water	NA	NA	20.7	7.01	01/07/2022	13:35
5	CCV1	1	Water	NA	NA	20.7	2.02	01/07/2022	13:36
6	N1073-02	1	Water	NA	NA	20.1	5.60	01/07/2022	13:40
7	N1073-02DUP	1	Water	NA	NA	20.2	5.61	01/07/2022	13:41
8	CCV2	1	Water	NA	NA	20.7	12.03	01/07/2022	13:45

118115

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n1057 w

WorkList ID : 155846

Department : Wet-Chemistry

Date : 01-07-2022 13:15:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2022	Water	N1073-02 -D	pH	Cool 4 deg C	TETR06	H11	TT-RW7B-IDWGW-2022C	01/06/2022	9040C

Date/Time 01/07/22 13:20
Raw Sample Received by: 70 (09/)
Raw Sample Relinquished by: CP gm

Date/Time 01/07/22 14:30
Raw Sample Received by: CP gm
Raw Sample Relinquished by: 70 (09/)