

LB118117


7470A

INSTRUMENT ID: CV1

Linear

$\mu$  Abs.:

4049E



A= 0.0000e+000

B= 2.5290e-004

C= 4.7178e-003

Rho= 0.9987809

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.042  | 0.042  | 146   | 0.000      | 146   |       |       |       |       | -  |
| 0.2    | 0.200  | 0.219  | 0.019  | 846   | 0.0 %      | 846   |       |       |       |       | 9  |
| 2.5    | 2.500  | 2.644  | 0.144  | 10437 | 0.0 %      | 10437 |       |       |       |       | 6  |
| 5.0    | 5.000  | 4.858  | -0.142 | 19190 | 0.0 %      | 19190 |       |       |       |       | -3 |
| 7.5    | 7.500  | 7.191  | -0.309 | 28416 | 0.0 %      | 28416 |       |       |       |       | -4 |
| 10.0   | 10.000 | 10.246 | 0.246  | 40496 | 0.0 %      | 40496 |       |       |       |       | 2  |

## LB118117 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jan 2022 16:08:05

| Sample ID       | Extended ID               | μ Abs. | Conc.   | Std ConcMethod | Units | Date                 | Type | Type |
|-----------------|---------------------------|--------|---------|----------------|-------|----------------------|------|------|
| 0.0 - 1         | 50                        | 146    | -       | 0.00007470A    | PPB   | 07 Jan 2022 16:10:08 | S    | Std  |
| 0.2 - 1         | 50.2                      | 846    | -       | 0.20007470A    | PPB   | 07 Jan 2022 16:12:25 | S    | Std  |
| 2.5 - 1         | 52.5                      | 10437  | -       | 2.50007470A    | PPB   | 07 Jan 2022 16:14:42 | S    | Std  |
| 5.0 - 1         | 55                        | 19190  | -       | 5.00007470A    | PPB   | 07 Jan 2022 16:16:59 | S    | Std  |
| 7.5 - 1         | 57.5                      | 28416  | -       | 7.50007470A    | PPB   | 07 Jan 2022 16:25:58 | S    | Std  |
| 10.0 - 1        | 516                       | 40496  | -       | 10.00007470A   | PPB   | 07 Jan 2022 16:30:44 | S    | Std  |
| ICV36 - 1       | ICV36                     | 15790  | 3.9981  | -7470A         | PPB   | 07 Jan 2022 16:35:59 | U    | SMPL |
| ICB36 - 1       | ICB36                     | -25    | -0.0016 | -7470A         | PPB   | 07 Jan 2022 16:38:14 | U    | SMPL |
| CCV58 - 1       | CCV58                     | 19147  | 4.8471  | -7470A         | PPB   | 07 Jan 2022 16:40:29 | U    | SMPL |
| CCB58 - 1       | CCB58                     | -33    | -0.0036 | -7470A         | PPB   | 07 Jan 2022 16:42:45 | U    | SMPL |
| CRA - 1         | CRA                       | 780    | 0.2020  | -7470A         | PPB   | 07 Jan 2022 16:45:00 | U    | SMPL |
| HighStd - 1     | HighStd                   | 40996  | 10.3727 | -7470A         | PPB   | 07 Jan 2022 16:47:15 | U    | SMPL |
| ChkStd - 1      | ChkStd                    | 27158  | 6.8731  | -7470A         | PPB   | 07 Jan 2022 16:49:30 | U    | SMPL |
| PB141922BL - 1  | PBW                       | 50     | 0.0174  | -7470A         | PPB   | 07 Jan 2022 16:51:47 | U    | SMPL |
| PB141922BS - 1  | LCSW                      | 13498  | 3.4184  | -7470A         | PPB   | 07 Jan 2022 16:54:03 | U    | SMPL |
| N1073-02 - 1    | TT-RW7B-IDWGW-20220106    | 45     | 0.0161  | -7470A         | PPB   | 07 Jan 2022 16:56:19 | U    | SMPL |
| N1073-02DUP - 1 | TT-RW7B-IDWGW-20220106DUP | 88     | 0.0270  | -7470A         | PPB   | 07 Jan 2022 16:58:35 | U    | SMPL |
| N1073-02MS - 1  | TT-RW7B-IDWGW-20220106MS  | 13188  | 3.3400  | -7470A         | PPB   | 07 Jan 2022 17:00:53 | U    | SMPL |
| N1073-02MSD - 1 | TT-RW7B-IDWGW-20220106MSD | 12741  | 3.2270  | -7470A         | PPB   | 07 Jan 2022 17:03:10 | U    | SMPL |
| CCV59 - 1       | CCV59                     | 19262  | 4.8761  | -7470A         | PPB   | 07 Jan 2022 17:05:25 | U    | SMPL |
| CCB59 - 1       | CCB59                     | -31    | -0.0031 | -7470A         | PPB   | 07 Jan 2022 17:07:41 | U    | SMPL |
| N1073-02LX5 - 1 |                           | 70     | 0.0224  | -7470A         | PPB   | 07 Jan 2022 17:09:56 | U    | SMPL |
| N1073-02A - 1   |                           | 15088  | 3.8205  | -7470A         | PPB   | 07 Jan 2022 17:12:11 | U    | SMPL |
| CCV60 - 1       | CCV60                     | 19004  | 4.8109  | -7470A         | PPB   | 07 Jan 2022 17:14:27 | U    | SMPL |
| CCB60 - 1       | CCB60                     | -38    | -0.0049 | -7470A         | PPB   | 07 Jan 2022 17:16:42 | U    | SMPL |