


LB118134

7470A

1 MICROMENT 10: CV



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.0	0.000	-0.024	-0.024	42	0.000	42					-11
0.2	0.200	0.178	-0.022	814	0.0 %	814					3
2.5	2.500	2.584	0.084	9981	0.0 %	9981					1
5.0	5.000	5.030	0.030	19301	0.0 %	19301					-2
7.5	7.500	7.379	-0.121	28252	0.0 %	28252					1
10.0	10.000	10.052	0.052	38435	0.0 %	38435					

LB118134 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 10 Jan 2022 10:50:54				
Sample ID	Extended ID	µ Abs	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	42	-	0.00007470A	PPB	10 Jan 2022 10:59:20	S	Std
0.2 - 1	50.2	814	-	0.20007470A	PPB	10 Jan 2022 11:01:38	S	Std
2.5 - 1	52.5	9981	-	2.50007470A	PPB	10 Jan 2022 11:03:55	S	Std
5.0 - 1	55	19301	-	5.00007470A	PPB	10 Jan 2022 11:06:12	S	Std
7.5 - 1	57.5	28252	-	7.50007470A	PPB	10 Jan 2022 11:08:29	S	Std
10.0 - 1	510	38435	-	10.00007470A	PPB	10 Jan 2022 11:13:17	S	Std
ICV37 - 1	ICV37	15789	4.1085	-7470A	PPB	10 Jan 2022 11:15:54	U	SMPL
ICB37 - 1	ICB37	-3	-0.0361	-7470A	PPB	10 Jan 2022 11:18:09	U	SMPL
CCV61 - 1	CCV61	20408	5.3208	-7470A	PPB	10 Jan 2022 11:20:24	U	SMPL
CCB61 - 1	CCB61	-3	-0.0361	-7470A	PPB	10 Jan 2022 11:22:39	U	SMPL
CRA - 1	CRA	750	0.1615	-7470A	PPB	10 Jan 2022 11:24:55	U	SMPL
HighStd - 1	HighStd	38274	10.0098	-7470A	PPB	10 Jan 2022 11:27:09	U	SMPL
ChkStd - 1	ChkStd	25587	6.6800	-7470A	PPB	10 Jan 2022 11:29:24	U	SMPL
PB141920BL - 1	PBW	17	-0.0309	-7470A	PPB	10 Jan 2022 11:31:40	U	SMPL
PB141920BS - 1	LCSW	14436	3.7534	-7470A	PPB	10 Jan 2022 11:33:56	U	SMPL
N1004-04 - 1	RO-R2-IEC	42	-0.0243	-7470A	PPB	10 Jan 2022 11:36:12	U	SMPL
N1052-02 - 1	BALL-1	84	-0.0133	-7470A	PPB	10 Jan 2022 11:38:29	U	SMPL
N1055-03 - 1	149TH-ST-STATION-DRUM-1-PPE	84	-0.0133	-7470A	PPB	10 Jan 2022 11:40:45	U	SMPL
N1055-04 - 1	149TH-ST-STATION-DRUM-3-POLY	205	0.0185	-7470A	PPB	10 Jan 2022 11:43:03	U	SMPL
N1056-01 - 1	COMP-1	233	0.0258	-7470A	PPB	10 Jan 2022 11:45:19	U	SMPL
CCV62 - 1	CCV62	20351	5.3058	-7470A	PPB	10 Jan 2022 11:47:34	U	SMPL
CCB62 - 1	CCB62	3	-0.0345	-7470A	PPB	10 Jan 2022 11:49:50	U	SMPL
N1056-02 - 1	COMP-2	203	0.0180	-7470A	PPB	10 Jan 2022 11:52:05	U	SMPL
N1056-03 - 1	COMP-4	204	0.0182	-7470A	PPB	10 Jan 2022 11:54:20	U	SMPL
N1056-04 - 1	1906	181	0.0122	-7470A	PPB	10 Jan 2022 11:56:35	U	SMPL
N1056-05 - 1	1896	184	0.0130	-7470A	PPB	10 Jan 2022 11:58:51	U	SMPL
N1056-06 - 1	1917-OIL	2425	0.6011	-7470A	PPB	10 Jan 2022 12:01:07	U	SMPL
N1067-02 - 1	124035	116	-0.0049	-7470A	PPB	10 Jan 2022 12:07:48	U	SMPL
N1067-02DUP - 1	124035DUP	133	-0.0004	-7470A	PPB	10 Jan 2022 12:10:04	U	SMPL
N1067-02MS - 1	124035MS	15829	4.1190	-7470A	PPB	10 Jan 2022 12:12:21	U	SMPL
N1067-02MSD - 1	124035MSD	14324	3.7240	-7470A	PPB	10 Jan 2022 12:14:38	U	SMPL
PB141921BL - 1	PBW	65	-0.0183	-7470A	PPB	10 Jan 2022 12:16:53	U	SMPL
CCV63 - 1	CCV63	20595	5.3699	-7470A	PPB	10 Jan 2022 12:19:09	U	SMPL
CCB63 - 1	CCB63	1	-0.0350	-7470A	PPB	10 Jan 2022 12:21:24	U	SMPL
PB141921BS - 1	LCSW	13940	3.6233	-7470A	PPB	10 Jan 2022 12:23:40	U	SMPL
N1056-07 - 1	COMP-3	197	0.0164	-7470A	PPB	10 Jan 2022 12:25:56	U	SMPL
N1056-08 - 1	1917-WATER	27	-0.0282	-7470A	PPB	10 Jan 2022 12:28:12	U	SMPL
N1056-08DUP - 1	1917-WATERDUP	73	-0.0162	-7470A	PPB	10 Jan 2022 12:30:27	U	SMPL
N1056-08MS - 1	1917-WATERMS	13690	3.5576	-7470A	PPB	10 Jan 2022 12:32:44	U	SMPL
N1056-08MSD - 1	1917-WATERMSD	13244	3.4406	-7470A	PPB	10 Jan 2022 12:35:00	U	SMPL
PB141923BL - 1	PBW	2	-0.0348	-7470A	PPB	10 Jan 2022 12:37:16	U	SMPL
PB141923BS - 1	LCSW	16030	4.1718	-7470A	PPB	10 Jan 2022 12:39:33	U	SMPL
N1050-03 - 1	P1660	4142	1.0518	-7470A	PPB	10 Jan 2022 12:41:50	U	SMPL
N1050-03DUP - 1	P1660DUP	4326	1.1001	-7470A	PPB	10 Jan 2022 12:44:06	U	SMPL
CCV64 - 1	CCV64	20442	5.3297	-7470A	PPB	10 Jan 2022 12:46:21	U	SMPL
CCB64 - 1	CCB64	18	-0.0306	-7470A	PPB	10 Jan 2022 12:48:37	U	SMPL
N1050-03MS - 1	P1660MS	18000	4.6888	-7470A	PPB	10 Jan 2022 12:50:52	U	SMPL
N1050-03MSD - 1	P1660MSD	18150	4.7282	-7470A	PPB	10 Jan 2022 12:53:08	U	SMPL
N1054-07 - 1	20220019-CAPE-1-MODIFIED-ELUTRIATE	101	-0.0088	-7470A	PPB	10 Jan 2022 12:55:23	U	SMPL
N1054-09 - 1	20220020-CAPE-2-MODIFIED-ELUTRIATE	146	0.0030	-7470A	PPB	10 Jan 2022 12:57:39	U	SMPL
N1054-11 - 1	20220021-CAPE-3-MODIFIED-ELUTRIATE	95	-0.0104	-7470A	PPB	10 Jan 2022 12:59:54	U	SMPL
N1054-13 - 1	20220022-CAPE-4-MODIFIED-ELUTRIATE	77	-0.0151	-7470A	PPB	10 Jan 2022 13:02:11	U	SMPL
N1054-15 - 1	20220023-CAPE-5-MODIFIED-ELUTRIATE	99	-0.0093	-7470A	PPB	10 Jan 2022 13:04:27	U	SMPL
N1054-17 - 1	20220024-CAPE-6-MODIFIED-ELUTRIATE	87	-0.0125	-7470A	PPB	10 Jan 2022 13:06:44	U	SMPL
N1071-04 - 1	FB-010622	18	-0.0306	-7470A	PPB	10 Jan 2022 13:09:01	U	SMPL
N1078-03 - 1	TEMP-WELL-1	353	0.0573	-7470A	PPB	10 Jan 2022 13:11:17	U	SMPL
CCV65 - 1	CCV65	18767	4.8901	-7470A	PPB	10 Jan 2022 13:13:33	U	SMPL
CCB65 - 1	CCB65	45	-0.0235	-7470A	PPB	10 Jan 2022 13:15:48	U	SMPL
PB141914TB - 1	PB141914TB	79	-0.0146	-7470A	PPB	10 Jan 2022 13:18:04	U	SMPL
PB141918TB - 1	PB141918TB	174	0.0104	-7470A	PPB	10 Jan 2022 13:20:20	U	SMPL
N1067-02LX5 - 1		0	-0.0353	-7470A	PPB	10 Jan 2022 13:22:35	U	SMPL
N1067-02A - 1		15675	4.0786	-7470A	PPB	10 Jan 2022 13:24:51	U	SMPL
N1056-08LX5 - 1		276	0.0371	-7470A	PPB	10 Jan 2022 13:27:07	U	SMPL
N1056-08A - 1		15760	4.1009	-7470A	PPB	10 Jan 2022 13:29:24	U	SMPL
N1050-03LX5 - 1		780	0.1694	-7470A	PPB	10 Jan 2022 13:31:40	U	SMPL
N1050-03A - 1		19326	5.0368	-7470A	PPB	10 Jan 2022 13:33:57	U	SMPL
CCV66 - 1	CCV66	18941	4.9358	-7470A	PPB	10 Jan 2022 13:36:14	U	SMPL
CCB66 - 1	CCB66	-11	-0.0382	-7470A	PPB	10 Jan 2022 13:38:31	U	SMPL